

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042919\
 Data File : PL047989.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2019 19:32
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 01:32:07 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 25 01:53:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.941	479.3E6	127.0E6	51.830	53.185
28)	SA Decachlor...	7.998	8.977	497.0E6	129.9E6	44.396	45.510
Target Compounds							
2)	A alpha-BHC	3.757	4.329	773.4E6	180.5E6	52.909	54.389
3)	MA gamma-BHC...	4.033	4.613	711.1E6	167.0E6	53.646	53.249
4)	MA Heptachlor	4.318	5.116	716.8E6	169.2E6	50.306	50.334
5)	MB Aldrin	4.557	5.419	713.4E6	161.9E6	51.226	51.863
6)	B beta-BHC	4.286	4.786	288.3E6	77368632	51.035	53.697
7)	B delta-BHC	4.481	5.001	681.9E6	150.7E6	52.492	59.972
8)	B Heptachlor...	4.993	5.805	640.4E6	149.7E6	49.974	53.548
9)	A Endosulfan I	5.324	6.160	601.6E6	139.5E6	49.750	49.665
10)	B gamma-Chl...	5.216	6.039	664.1E6	153.3E6	49.883	50.639
11)	B alpha-Chl...	5.272	6.111	647.5E6	151.3E6	49.618	50.002
12)	B 4,4'-DDE	5.440	6.270	606.4E6	139.3E6	48.753	50.430
13)	MA Dieldrin	5.564	6.417	644.5E6	145.7E6	49.644	49.845
14)	MA Endrin	5.816	6.634	534.1E6	114.4E6	44.943	44.800
15)	B Endosulfa...	6.091	6.845	537.6E6	126.0E6	48.015	48.020
16)	A 4,4'-DDD	5.951	6.760	511.9E6	117.1E6	52.061	52.138
17)	MA 4,4'-DDT	6.185	7.057	451.9E6	104.6E6	45.322	43.900
18)	B Endrin al...	6.258	6.969	448.9E6	110.1E6	47.580	47.668
19)	B Endosulfa...	6.469	7.191	534.1E6	128.7E6	47.916	46.983
20)	A Methoxychlor	6.730	7.517	232.5E6	60169522	43.717	42.912
21)	B Endrin ke...	6.951	7.660	590.6E6	136.4E6	49.801	48.500
22)	Mirex	7.126	8.110	427.4E6	104.9E6	45.530	45.099

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042919\
Data File : PL047989.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2019 19:32
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 01:32:07 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
Quant Title : GC Extractables
QLast Update : Thu Apr 25 01:53:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

