

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042921\
 Data File : PL066565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2021 16:01
 Operator : AR\AJ
 Sample : PB136061BS
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB136061BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 09:12:30 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 28 08:49:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.434	4.154	90408963	88078401	23.993	25.079
28)	SA Decachlor...	8.175	9.358	81301145	67024146	21.204	21.666
Target Compounds							
2)	A alpha-BHC	3.868	4.560	324.2E6	285.1E6	52.090	55.410
3)	MA gamma-BHC...	4.148	4.853	295.8E6	264.0E6	53.756	54.725
4)	MA Heptachlor	4.447	5.370	289.9E6	257.0E6	51.918	52.057
5)	MB Aldrin	4.691	5.681	280.7E6	236.9E6	53.825	50.676
6)	B beta-BHC	4.397	5.029	126.0E6	121.4E6	47.626	52.753
7)	B delta-BHC	4.598	5.253	301.9E6	254.0E6	53.058	57.340
8)	B Heptachlo...	5.131	6.076	258.6E6	211.8E6	52.815	57.246
9)	A Endosulfan I	5.466	6.441	240.2E6	193.8E6	52.055	50.888
10)	B gamma-Chl...	5.357	6.313	266.2E6	215.5E6	52.248	51.762
11)	B alpha-Chl...	5.415	6.386	258.0E6	211.1E6	52.402	50.889
12)	B 4,4'-DDE	5.583	6.542	239.6E6	196.1E6	52.505	51.688
13)	MA Dieldrin	5.708	6.702	251.2E6	200.4E6	51.777	51.175
14)	MA Endrin	5.962	6.925	197.9E6	135.8E6	45.440	43.324
15)	B Endosulfa...	6.235	7.139	206.7E6	161.3E6	50.094	48.840
16)	A 4,4'-DDD	6.097	7.042	211.3E6	166.4E6	55.642	53.414
17)	MA 4,4'-DDT	6.336	7.343	179.7E6	140.6E6	46.076	43.436
18)	B Endrin al...	6.403	7.265	166.8E6	131.5E6	49.240	47.019
19)	B Endosulfa...	6.616	7.489	214.5E6	157.9E6	49.630	46.172
20)	A Methoxychlor	6.887	7.803	92330655	71054787	47.005	39.612
21)	B Endrin ke...	7.102	7.970	236.1E6	172.3E6	53.894	48.556
22)	Mirex	7.289	8.425	169.7E6	122.1E6	46.190	43.770

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042921\
Data File : PL066565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2021 16:01
Operator : AR\AJ
Sample : PB136061BS
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB136061BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 09:12:30 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
Quant Title : GC Extractables
QLast Update : Wed Apr 28 08:49:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

