

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042921\
 Data File : PL066585.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2021 22:10
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 04:16:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 28 08:49:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.430	4.155	201.3E6	201.6E6	53.422	57.412
28)	SA Decachlor...	8.168	9.357	179.3E6	156.2E6	46.771	50.487
Target Compounds							
2)	A alpha-BHC	3.863	4.559	340.3E6	308.1E6	54.681	59.892
3)	MA gamma-BHC...	4.143	4.853	307.9E6	282.3E6	55.955	58.515
4)	MA Heptachlor	4.441	5.369	313.6E6	285.5E6	56.160	57.829
5)	MB Aldrin	4.686	5.681	294.0E6	257.4E6	56.378	55.062
6)	B beta-BHC	4.392	5.029	135.9E6	134.9E6	51.366	58.951
7)	B delta-BHC	4.593	5.253	320.9E6	270.7E6	56.407	61.117
8)	B Heptachlo...	5.125	6.076	273.6E6	210.5E6	55.861	56.910
9)	A Endosulfan I	5.459	6.441	256.8E6	216.8E6	55.653	56.940
10)	B gamma-Chl...	5.350	6.313	282.2E6	240.4E6	55.369	57.750
11)	B alpha-Chl...	5.409	6.386	274.8E6	236.1E6	55.800	56.921
12)	B 4,4'-DDE	5.576	6.541	261.1E6	222.3E6	57.218	58.584
13)	MA Dieldrin	5.701	6.702	275.1E6	227.0E6	56.708	57.982
14)	MA Endrin	5.955	6.925	247.2E6	183.4E6	56.758	58.533
15)	B Endosulfa...	6.227	7.139	228.8E6	184.4E6	55.439	55.845
16)	A 4,4'-DDD	6.088	7.041	217.0E6	178.2E6	57.123	57.196
17)	MA 4,4'-DDT	6.328	7.343	224.9E6	184.4E6	57.676	56.967
18)	B Endrin al...	6.395	7.265	185.6E6	153.4E6	54.777	54.832
19)	B Endosulfa...	6.608	7.489	239.5E6	186.5E6	55.416	54.514
20)	A Methoxychlor	6.877	7.803	117.4E6	96309163	59.791	53.690
21)	B Endrin ke...	7.094	7.970	243.8E6	188.9E6	55.649	53.215
22)	Mirex	7.282	8.426	196.5E6	141.8E6	53.500	50.834

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042921\
Data File : PL066585.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2021 22:10
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 04:16:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
Quant Title : GC Extractables
QLast Update : Wed Apr 28 08:49:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

