

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042921\
 Data File : PL066595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2021 01:46
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 09:38:24 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 28 08:49:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.429	4.155	206.3E6	211.1E6	54.762	60.116
28)	SA Decachlor...	8.165	9.357	179.3E6	154.7E6	46.756	50.016
Target Compounds							
2)	A alpha-BHC	3.863	4.560	350.2E6	322.7E6	56.267	62.729
3)	MA gamma-BHC...	4.143	4.853	314.8E6	294.9E6	57.222	61.123
4)	MA Heptachlor	4.441	5.370	316.9E6	297.2E6	56.757	60.193
5)	MB Aldrin	4.686	5.682	297.2E6	271.3E6	56.992	58.033
6)	B beta-BHC	4.392	5.029	138.4E6	142.2E6	52.307	62.326
7)	B delta-BHC	4.592	5.253	325.9E6	284.0E6	57.287	64.103
8)	B Heptachlo...	5.125	6.076	275.6E6	226.6E6	56.282	61.255
9)	A Endosulfan I	5.460	6.441	257.6E6	223.7E6	55.830	58.742
10)	B gamma-Chl...	5.351	6.314	284.9E6	252.6E6	55.902	60.684
11)	B alpha-Chl...	5.408	6.387	276.5E6	244.8E6	56.164	59.020
12)	B 4,4'-DDE	5.576	6.542	260.8E6	226.9E6	57.167	59.789
13)	MA Dieldrin	5.701	6.703	280.5E6	230.5E6	57.821	58.879
14)	MA Endrin	5.955	6.926	247.1E6	182.7E6	56.714	58.302
15)	B Endosulfa...	6.227	7.139	229.5E6	187.5E6	55.619	56.760
16)	A 4,4'-DDD	6.089	7.042	219.1E6	182.0E6	57.676	58.423
17)	MA 4,4'-DDT	6.328	7.343	223.8E6	184.0E6	57.393	56.847
18)	B Endrin al...	6.395	7.265	186.2E6	154.8E6	54.959	55.341
19)	B Endosulfa...	6.607	7.490	240.4E6	190.1E6	55.624	55.582
20)	A Methoxychlor	6.878	7.804	116.8E6	96381313	59.449	53.731
21)	B Endrin ke...	7.094	7.970	241.0E6	192.3E6	55.017	54.170
22)	Mirex	7.282	8.426	195.6E6	142.2E6	53.245	50.974

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL042921\
Data File : PL066595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2021 01:46
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 09:38:24 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
Quant Title : GC Extractables
QLast Update : Wed Apr 28 08:49:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

