

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL043019\
 Data File : PL048019.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Apr 2019 12:21
 Operator : AJ\SJ
 Sample : K2571-01MSD
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 114-PENNINGTONMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 01:07:59 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 25 01:53:03 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.940	137.9E6	44024620	14.909	18.441
28)	SA Decachlor...	7.997	8.976	175.1E6	47644912	15.645	16.691
Target Compounds							
2)	A alpha-BHC	3.756	4.329	555.8E6	142.9E6	38.021	43.056
3)	MA gamma-BHC...	4.033	4.612	489.0E6	132.4E6	36.892	42.210
4)	MA Heptachlor	4.318	5.116	525.1E6	138.0E6	36.853	41.058
5)	MB Aldrin	4.557	5.418	510.3E6	123.7E6	36.647	39.626
6)	B beta-BHC	4.286	4.786	195.4E6	59235813	34.583	41.112
7)	B delta-BHC	4.480	5.000	390.9E6	99597131	30.093	39.644 #
8)	B Heptachlor...	4.993	5.805	453.9E6	120.7E6	35.417	43.182
9)	A Endosulfan I	5.324	6.161	439.9E6	107.8E6	36.376	38.397
10)	B gamma-Chl...	5.216	6.038	480.9E6	126.2E6	36.120	41.704
11)	B alpha-Chl...	5.272	6.111	464.8E6	124.4E6	35.621	41.123
12)	B 4,4'-DDE	5.440	6.270	445.2E6	111.5E6	35.793	40.382
13)	MA Dieldrin	5.563	6.417	487.4E6	120.4E6	37.544	41.169
14)	MA Endrin	5.816	6.633	437.9E6	106.3E6	36.845	41.623
15)	B Endosulfa...	6.090	6.844	422.7E6	111.0E6	37.749	42.290
16)	A 4,4'-DDD	5.950	6.760	360.0E6	97348114	36.612	43.361
17)	MA 4,4'-DDT	6.185	7.057	381.8E6	103.7E6	38.292	43.527
18)	B Endrin al...	6.257	6.968	345.1E6	94018879	36.572	40.695
19)	B Endosulfa...	6.468	7.191	421.6E6	110.9E6	37.820	40.501
20)	A Methoxychlor	6.730	7.517	220.2E6	62819036	41.401	44.802
21)	B Endrin ke...	6.951	7.659	492.5E6	128.9E6	41.533	45.861
22)	Mirex	7.125	8.110	364.6E6	103.2E6	38.844	44.356

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL043019\
Data File : PL048019.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Apr 2019 12:21
Operator : AJ\SJ
Sample : K2571-01MSD
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
114-PENNINGTONMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 01:07:59 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042419.M
Quant Title : GC Extractables
QLast Update : Thu Apr 25 01:53:03 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

