

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050119\
 Data File : PL048088.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 May 2019 20:17
 Operator : AJ\SJ
 Sample : PCHLORICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 09:30:36 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:29:17 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.940	41383980	12156346	3.793	4.680
28)	SA Decachlor...	7.996	8.976	63086549	14206180	6.119	5.310
Target Compounds							
23)	Chlordane-1	4.169	4.927	17074888	10381358	35.697	82.458 #
24)	Chlordane-2	4.660	5.394	26441189	6382968	50.084	48.302
25)	Chlordane-3	5.216	6.038	66054142	20551255	39.272	45.886
26)	Chlordane-4	5.270	6.111	61352909	25550991	40.226	48.123
27)	Chlordane-5	6.093	6.913	27982522	4716558	58.034	45.237



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050119\
Data File : PL048088.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 01 May 2019 20:17
Operator : AJ\SJ
Sample : PCHLORICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 09:30:36 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
Quant Title : GC Extractables
QLast Update : Thu May 02 09:29:17 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.328 min
Delta R.T.: 0.000 min
Response: 41383980
Conc: 3.79 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.940 min
Delta R.T.: 0.000 min
Response: 12156346
Conc: 4.68 ng/ml




#23 Chlordane-1

R.T.: 4.169 min
Delta R.T.: 0.000 min
Response: 17074888
Conc: 35.70 ng/ml



#23 Chlordane-1


R.T.: 4.927 min
Delta R.T.: 0.000 min
Response: 10381358
Conc: 82.46 ng/ml



#24 Chlordane-2


R.T.: 4.660 min
Delta R.T.: 0.000 min
Response: 26441189
Conc: 50.08 ng/ml

Instrument :
ECD_L
ClientSampleId :




#24 Chlordane-2

R.T.: 5.394 min
Delta R.T.: 0.000 min
Response: 6382968
Conc: 48.30 ng/ml




#25 Chlordane-3

R.T.: 5.216 min
Delta R.T.: 0.000 min
Response: 66054142
Conc: 39.27 ng/ml



#25 Chlordane-3


R.T.: 6.038 min
Delta R.T.: 0.000 min
Response: 20551255
Conc: 45.89 ng/ml



#26 Chlordane-4


R.T.: 5.270 min
Delta R.T.: 0.000 min
Response: 61352909
Conc: 40.23 ng/ml

Instrument :
ECD_L
ClientSampleId :




#26 Chlordane-4

R.T.: 6.111 min
Delta R.T.: 0.000 min
Response: 25550991
Conc: 48.12 ng/ml




#27 Chlordane-5

R.T.: 6.093 min
Delta R.T.: 0.000 min
Response: 27982522
Conc: 58.03 ng/ml



#27 Chlordane-5


R.T.: 6.913 min
Delta R.T.: 0.000 min
Response: 4716558
Conc: 45.24 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.996 min
Delta R.T.: 0.000 min
Response: 63086549
Conc: 6.12 ng/ml

Instrument :
ECD_L
ClientSampleId :



#28 Decachlorobiphenyl

R.T.: 8.976 min
Delta R.T.: 0.000 min
Response: 14206180
Conc: 5.31 ng/ml