

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX050119.M
 Title : GC Extractables
 Last Update : Thu May 02 04:41:08 2019
 Response Via : Initial Calibration

Calibration Files

500 =PL048091.D 1000 =PL048089.D 250 =PL048092.D
 750 =PL048090.D 100 =PL048093.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	9.184	8.909	9.261	9.332	9.063	9.150	E6	1.83
2)	Toxaphene-1	7.094	6.639	6.247	6.996	6.306	6.656	E4	5.80
3)	Toxaphene-2	1.724	1.715	1.674	1.802	1.639	1.711	E5	3.59
4)	Toxaphene-3	8.260	7.887	7.617	8.142	6.945	7.770	E4	6.74
5)	Toxaphene-4	1.336	1.368	1.254	1.342	1.070	1.274	E5	9.55
6)	Toxaphene-5	2.231	2.159	2.137	2.247	2.065	2.168	E5	3.42
7) SA	Decachlorobiphenyl	0.961	0.926	1.007	0.967	1.086	0.989	E7	6.17

Signal #2 Calibration Files

500 =PL048091.D 1000 =PL048089.D 250 =PL048092.D
 750 =PL048090.D 100 =PL048093.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.396	2.324	2.416	2.437	2.427	2.400	E6	1.88
2)	Toxaphene-1	1.497	1.456	1.531	1.490	1.556	1.506	E4	2.54
3)	Toxaphene-2	1.818	1.744	1.664	1.704	1.841	1.754	E4	4.27
4)	Toxaphene-3	6.835	6.519	6.895	6.794	7.482	6.905	E4	5.12
5)	Toxaphene-4	4.206	4.135	4.098	4.183	4.168	4.158	E4	1.01
6)	Toxaphene-5	3.381	3.150	3.414	3.364	3.389	3.340	E4	3.22
7) SA	Decachlorobiphenyl	2.561	2.430	2.707	2.508	2.910	2.624	E6	7.23

(#) = Out of Range