

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050319\
 Data File : PL048175.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 May 2019 16:26
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 23:04:50 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:36:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.331	3.941	530.8E6	137.2E6	54.793	55.084
28)	SA Decachlor...	8.001	8.980	417.4E6	122.6E6	41.519	46.188
Target Compounds							
2)	A alpha-BHC	3.759	4.330	871.0E6	200.8E6	56.312	57.168
3)	MA gamma-BHC...	4.035	4.613	794.8E6	182.6E6	55.935	55.912
4)	MA Heptachlor	4.321	5.117	793.2E6	183.9E6	53.914	53.114
5)	MB Aldrin	4.559	5.420	738.9E6	165.2E6	53.928	53.333
6)	B beta-BHC	4.289	4.788	317.0E6	80146820	53.712	51.508
7)	B delta-BHC	4.483	5.002	781.2E6	155.6E6	55.333	55.032
8)	B Heptachlo...	4.996	5.806	593.4E6	152.9E6	46.056	52.412
9)	A Endosulfan I	5.327	6.161	568.7E6	140.8E6	47.271	49.783
10)	B gamma-Chl...	5.219	6.040	638.2E6	154.7E6	47.796	50.249
11)	B alpha-Chl...	5.275	6.112	613.5E6	151.8E6	46.927	49.459
12)	B 4,4'-DDE	5.443	6.272	533.9E6	136.3E6	42.087	49.359
13)	MA Dieldrin	5.567	6.418	643.7E6	144.9E6	49.437	49.390
14)	MA Endrin	5.819	6.635	559.6E6	123.7E6	45.968	47.436
15)	B Endosulfa...	6.094	6.846	497.3E6	128.3E6	43.089	47.580
16)	A 4,4'-DDD	5.954	6.762	463.6E6	113.3E6	47.629	50.477
17)	MA 4,4'-DDT	6.189	7.059	412.0E6	104.5E6	41.005	44.550
18)	B Endrin al...	6.261	6.970	405.8E6	108.6E6	41.760	46.337
19)	B Endosulfa...	6.472	7.193	462.7E6	119.3E6	43.112	46.091
20)	A Methoxychlor	6.734	7.520	220.3E6	64595143	40.463	47.316
21)	B Endrin ke...	6.954	7.662	517.9E6	124.9E6	44.436	47.794
22)	Mirex	7.129	8.112	344.2E6	95848182	38.398	45.450

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050319\
Data File : PL048175.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 May 2019 16:26
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 23:04:50 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
Quant Title : GC Extractables
QLast Update : Thu May 02 09:36:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

