

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050421\
 Data File : PL066654.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2021 10:34
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 12:50:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 28 08:49:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.429	4.153	182.5E6	171.0E6	48.439	48.679
28)	SA Decachlor...	8.163	9.352	171.2E6	135.1E6	44.641	43.671
Target Compounds							
2)	A alpha-BHC	3.863	4.558	318.2E6	264.8E6	51.137	51.480
3)	MA gamma-BHC...	4.143	4.851	282.5E6	241.3E6	51.339	50.030
4)	MA Heptachlor	4.440	5.367	289.2E6	236.3E6	51.798	47.864
5)	MB Aldrin	4.685	5.679	274.3E6	216.8E6	52.585	46.365
6)	B beta-BHC	4.392	5.028	125.4E6	125.7E6	47.393	54.708
7)	B delta-BHC	4.592	5.251	293.4E6	222.9E6	51.566	50.325
8)	B Heptachlo...	5.124	6.074	253.5E6	173.3E6	51.771	46.845
9)	A Endosulfan I	5.458	6.438	239.7E6	177.4E6	51.954	46.591
10)	B gamma-Chl...	5.350	6.310	264.1E6	194.3E6	51.823	46.686
11)	B alpha-Chl...	5.407	6.383	257.1E6	191.8E6	52.219	46.233
12)	B 4,4'-DDE	5.575	6.538	242.8E6	179.7E6	53.211	47.368
13)	MA Dieldrin	5.699	6.699	254.9E6	182.9E6	52.540	46.709
14)	MA Endrin	5.954	6.922	227.2E6	147.2E6	52.163	46.979
15)	B Endosulfa...	6.226	7.136	210.1E6	147.4E6	50.918	44.627
16)	A 4,4'-DDD	6.087	7.039	201.7E6	142.1E6	53.099	45.627
17)	MA 4,4'-DDT	6.327	7.340	197.2E6	131.6E6	50.565	40.671
18)	B Endrin al...	6.394	7.261	173.0E6	124.8E6	51.060	44.629
19)	B Endosulfa...	6.607	7.486	220.4E6	151.4E6	51.000	44.262
20)	A Methoxychlor	6.876	7.800	107.4E6	72875170	54.678	40.627 #
21)	B Endrin ke...	7.092	7.967	220.6E6	152.8E6	50.344	43.043
22)	Mirex	7.279	8.421	178.0E6	116.2E6	48.453	41.646

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050421\
Data File : PL066654.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2021 10:34
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 12:50:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
Quant Title : GC Extractables
QLast Update : Wed Apr 28 08:49:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

