

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050422\
 Data File : PL074526.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2022 18:47
 Operator : AR\AJ
 Sample : N2665-10
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SB-10,8

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/05/2022
 Supervised By : mohammad ahmed 05/05/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 23:29:10 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
 Quant Title : GC Extractables
 QLast Update : Thu Apr 14 09:29:23 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.633	5.489	7275978	39076623	22.519	21.917
2) SA Decachlor...	10.402	11.333	10818028	37869537	20.803m	21.440

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050422\
Data File : PL074526.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2022 18:47
Operator : AR\AJ
Sample : N2665-10
Misc :
ALS Vial : 20 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
SB-10.8


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/05/2022
Supervised By : mohammad ahmed 05/05/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 23:29:10 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL040822.M
Quant Title : GC Extractables
QLast Update : Thu Apr 14 09:29:23 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 4.633 min
Delta R.T.: -0.006 min
Response: 7275978
Conc: 22.52 ng/ml

Instrument :
ECD_L
ClientSampleId :
SB-10,8


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/05/2022
Supervised By :mohammad ahmed 05/05/2022




#1 Tetrachloro-m-xylene

R.T.: 5.489 min
Delta R.T.: -0.002 min
Response: 39076623
Conc: 21.92 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.402 min
Delta R.T.: -0.012 min
Response: 10818028
Conc: 20.80 ng/ml m



#28 Decachlorobiphenyl

R.T.: 11.333 min
Delta R.T.: -0.003 min
Response: 37869537
Conc: 21.44 ng/ml