

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050521\
 Data File : PL066703.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2021 16:05
 Operator : AR\AJ
 Sample : M2277-01MSD
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 FK-BPM-01-A3-A3A-B3-B3AMSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 06 07:48:05 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 28 08:49:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.426	4.151	52642568	52167478	13.971	14.854
28)	SA Decachlor...	8.143	9.347	59524562	41507049	15.525m	13.417
Target Compounds							
2)	A alpha-BHC	3.859	4.556	354.1E6	304.2E6	56.900	59.123
3)	MA gamma-BHC...	4.138	4.849	312.2E6	279.4E6	56.736	57.925
4)	MA Heptachlor	4.436	5.365	318.6E6	276.0E6	57.064	55.910
5)	MB Aldrin	4.680	5.677	301.9E6	255.4E6	57.882	54.627
6)	B beta-BHC	4.387	5.026	132.3E6	141.8E6	50.006	62.139
7)	B delta-BHC	4.587	5.249	323.6E6	268.8E6	56.882	60.687
8)	B Heptachlo...	5.117	6.071	280.8E6	218.6E6	57.334	59.100
9)	A Endosulfan I	5.451	6.435	263.7E6	204.1E6	57.165	53.604
10)	B gamma-Chl...	5.343	6.308	301.5E6	228.2E6	59.167	54.829
11)	B alpha-Chl...	5.400	6.380	290.7E6	228.4E6	59.048	55.056
12)	B 4,4'-DDE	5.567	6.537	268.8E6	208.8E6	58.907	55.031
13)	MA Dieldrin	5.692	6.697	280.9E6	213.8E6	57.896	54.603
14)	MA Endrin	5.944	6.919	246.4E6	168.5E6	56.575m	53.782
15)	B Endosulfa...	6.217	7.134	231.7E6	174.2E6	56.159	52.729
16)	A 4,4'-DDD	6.078	7.037	217.2E6	171.7E6	57.185	55.119
17)	MA 4,4'-DDT	6.317	7.337	215.8E6	160.8E6	55.352	49.691
18)	B Endrin al...	6.384	7.259	185.1E6	138.8E6	54.645	49.607
19)	B Endosulfa...	6.596	7.483	241.5E6	170.3E6	55.879	49.781
20)	A Methoxychlor	6.863	7.797	115.7E6	82775441	58.927	46.146
21)	B Endrin ke...	7.080	7.964	254.7E6	172.5E6	58.129m	48.594
22)	Mirex	7.268	8.418	199.9E6	130.1E6	54.428	46.650

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050521\
Data File : PL066703.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2021 16:05
Operator : AR\AJ
Sample : M2277-01MSD
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
FK-BPM-01-A3-A3A-B3-B3AMSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 06 07:48:05 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
Quant Title : GC Extractables
QLast Update : Wed Apr 28 08:49:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

