

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050619\
 Data File : PL048233.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2019 20:05
 Operator : AJ\SJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 00:51:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:36:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.941	502.5E6	130.5E6	51.871	52.383
28)	SA Decachlor...	7.997	8.978	482.2E6	127.6E6	47.957	48.073
Target Compounds							
2)	A alpha-BHC	3.756	4.329	829.2E6	188.7E6	53.607	53.717
3)	MA gamma-BHC...	4.032	4.613	755.8E6	174.0E6	53.194	53.263
4)	MA Heptachlor	4.317	5.116	767.1E6	177.8E6	52.140	51.366
5)	MB Aldrin	4.556	5.419	721.7E6	162.0E6	52.672	52.324
6)	B beta-BHC	4.286	4.787	304.9E6	79357117	51.660	51.000
7)	B delta-BHC	4.480	5.001	755.1E6	160.7E6	53.481	56.812
8)	B Heptachlor...	4.993	5.805	660.7E6	155.1E6	51.281	53.193
9)	A Endosulfan I	5.324	6.161	609.7E6	142.7E6	50.685	50.483
10)	B gamma-Chl...	5.216	6.040	686.6E6	156.7E6	51.422	50.897
11)	B alpha-Chl...	5.272	6.111	666.8E6	154.9E6	51.000	50.498
12)	B 4,4'-DDE	5.440	6.271	607.8E6	140.9E6	47.912	51.031
13)	MA Dieldrin	5.563	6.417	687.2E6	149.4E6	52.775	50.930
14)	MA Endrin	5.815	6.634	593.4E6	126.4E6	48.743	48.490
15)	B Endosulfa...	6.090	6.845	571.8E6	132.5E6	49.545	49.119
16)	A 4,4'-DDD	5.950	6.761	514.9E6	117.1E6	52.900	52.182
17)	MA 4,4'-DDT	6.185	7.058	485.5E6	111.5E6	48.329	47.523
18)	B Endrin al...	6.257	6.969	472.6E6	114.0E6	48.633	48.666
19)	B Endosulfa...	6.469	7.192	523.6E6	126.9E6	48.788	49.024
20)	A Methoxychlor	6.730	7.518	251.1E6	64698519	46.119	47.392
21)	B Endrin ke...	6.950	7.660	593.8E6	132.2E6	50.945	50.582
22)	Mirex	7.125	8.111	414.4E6	101.6E6	46.225	48.183

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050619\
Data File : PL048233.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2019 20:05
Operator : AJ\SJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 00:51:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
Quant Title : GC Extractables
QLast Update : Thu May 02 09:36:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

