

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050621\
 Data File : PL066747.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2021 19:03
 Operator : AR\AJ
 Sample : M2296-01MS
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 SP-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 05:58:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 28 08:49:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.426	4.151	72118398	70166815	19.139	19.979
28)	SA Decachlor...	8.144	9.346	70456189	51945041	18.376	16.791
Target Compounds							
2)	A alpha-BHC	3.858	4.556	303.0E6	254.8E6	48.688	49.523
3)	MA gamma-BHC...	4.138	4.848	266.8E6	234.2E6	48.494	48.547
4)	MA Heptachlor	4.435	5.364	271.6E6	230.4E6	48.647	46.670
5)	MB Aldrin	4.679	5.676	254.2E6	212.2E6	48.732	45.388
6)	B beta-BHC	4.386	5.025	115.5E6	119.4E6	43.650	51.827
7)	B delta-BHC	4.586	5.248	273.6E6	224.9E6	48.095	50.764
8)	B Heptachlo...	5.116	6.070	241.5E6	183.0E6	49.315	49.468
9)	A Endosulfan I	5.450	6.435	225.0E6	171.7E6	48.759	45.083
10)	B gamma-Chl...	5.341	6.307	254.7E6	189.7E6	49.973	45.569
11)	B alpha-Chl...	5.399	6.380	244.9E6	193.7E6	49.747	46.692
12)	B 4,4'-DDE	5.566	6.535	277.3E6	212.9E6	60.779	56.108
13)	MA Dieldrin	5.691	6.696	236.4E6	177.7E6	48.725	45.388
14)	MA Endrin	5.945	6.918	218.8E6	142.1E6	50.226	45.343
15)	B Endosulfa...	6.215	7.131	194.1E6	140.4E6	47.051	42.513m
16)	A 4,4'-DDD	6.077	7.035	186.7E6	154.1E6	49.144	49.465
17)	MA 4,4'-DDT	6.315	7.336	199.4E6	145.7E6	51.138	45.015
18)	B Endrin al...	6.383	7.258	157.4E6	117.2E6	46.452	41.886
19)	B Endosulfa...	6.595	7.482	208.0E6	145.3E6	48.135	42.487
20)	A Methoxychlor	6.862	7.796	98146467	69001379	49.965	38.467
21)	B Endrin ke...	7.080	7.963	221.0E6	149.1E6	50.452	42.011
22)	Mirex	7.267	8.417	171.5E6	114.3E6	46.681	40.956

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050621\
Data File : PL066747.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2021 19:03
Operator : AR\AJ
Sample : M2296-01MS
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
SP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 05:58:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
Quant Title : GC Extractables
QLast Update : Wed Apr 28 08:49:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

