

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050621\
 Data File : PL066751.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 May 2021 20:01
 Operator : AR\AJ
 Sample : PB136201BS
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB136201BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 07 02:20:38 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 28 08:49:49 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.425	4.150	92229820	90953944	24.477	25.898
28)	SA Decachlor...	8.144	9.345	95730302	71427833	24.967	23.089
Target Compounds							
2)	A alpha-BHC	3.858	4.555	327.3E6	283.0E6	52.596	55.007
3)	MA gamma-BHC...	4.137	4.848	293.8E6	257.2E6	53.393	53.308
4)	MA Heptachlor	4.435	5.363	291.2E6	245.5E6	52.164	49.731
5)	MB Aldrin	4.678	5.675	283.7E6	237.6E6	54.389	50.819
6)	B beta-BHC	4.386	5.024	127.0E6	130.6E6	48.003	56.975
7)	B delta-BHC	4.585	5.247	302.6E6	247.8E6	53.187	55.931
8)	B Heptachlo...	5.116	6.070	263.4E6	210.7E6	53.785	56.944
9)	A Endosulfan I	5.450	6.434	248.5E6	194.6E6	53.851	51.108
10)	B gamma-Chl...	5.341	6.306	272.9E6	212.3E6	53.560	51.011
11)	B alpha-Chl...	5.399	6.380	267.0E6	210.2E6	54.217	50.676
12)	B 4,4'-DDE	5.566	6.535	250.7E6	196.4E6	54.948	51.773
13)	MA Dieldrin	5.691	6.695	261.4E6	198.4E6	53.881	50.680
14)	MA Endrin	5.944	6.918	204.8E6	137.6E6	47.021	43.920
15)	B Endosulfa...	6.215	7.132	217.1E6	163.3E6	52.626	49.448
16)	A 4,4'-DDD	6.076	7.035	228.8E6	169.2E6	60.237	54.309
17)	MA 4,4'-DDT	6.315	7.336	180.2E6	135.0E6	46.208	41.720
18)	B Endrin al...	6.383	7.258	179.9E6	134.8E6	53.105	48.190
19)	B Endosulfa...	6.595	7.482	228.8E6	163.0E6	52.937	47.652
20)	A Methoxychlor	6.861	7.796	97261837	70553535	49.515	39.332
21)	B Endrin ke...	7.079	7.963	257.4E6	178.1E6	58.747	50.174
22)	Mirex	7.266	8.417	186.9E6	129.3E6	50.890	46.347

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050621\
Data File : PL066751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 May 2021 20:01
Operator : AR\AJ
Sample : PB136201BS
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB136201BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 07 02:20:38 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042621.M
Quant Title : GC Extractables
QLast Update : Wed Apr 28 08:49:49 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

