

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050720\
 Data File : PL058582.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 May 2020 19:16
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

mohammad
 5/8/2020 12:59:14 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 08 00:52:59 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042820.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 29 00:55:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.454	3.954	473.8E6	98152274	44.523	45.411
28)	SA Decachlor...	8.200	8.975	545.9E6	108.4E6	50.476	48.837
Target Compounds							
2)	A alpha-BHC	3.891	4.342	821.9E6	151.9E6	44.880	45.875
3)	MA gamma-BHC...	4.175	4.624	774.6E6	150.7E6	44.873	45.537
4)	MA Heptachlor	4.474	5.125	780.1E6	155.2E6	44.151	44.938
5)	MB Aldrin	4.721	5.427	745.2E6	150.9E6	44.927	44.656
6)	B beta-BHC	4.427	4.798	315.2E6	65510388	44.106	44.962
7)	B delta-BHC	4.630	5.011	762.5E6	145.9E6	44.882	58.030 #
8)	B Heptachlo...	5.163	5.812	687.3E6	155.3E6	45.253m	48.499
9)	A Endosulfan I	5.503	6.165	641.6E6	136.8E6	45.448	44.219
10)	B gamma-Chl...	5.390	6.043	692.1E6	148.9E6	44.945m	44.101m
11)	B alpha-Chl...	5.448	6.116	670.3E6	149.9E6	44.198m	44.086
12)	B 4,4'-DDE	5.614	6.276	676.7E6	131.1E6	44.754	42.984
13)	MA Dieldrin	5.746	6.422	676.0E6	135.7E6	45.341	43.371
14)	MA Endrin	6.003	6.637	606.8E6	113.4E6	44.639	44.126
15)	B Endosulfa...	6.276	6.849	556.6E6	109.8E6	43.990m	41.203
16)	A 4,4'-DDD	6.129	6.765	534.9E6	100.6E6	45.614	44.263
17)	MA 4,4'-DDT	6.369	7.061	510.9E6	90232057	43.786	41.209
18)	B Endrin al...	6.443	6.972	469.0E6	92284154	44.215m	42.327
19)	B Endosulfa...	6.658	7.195	512.7E6	100.2E6	44.206	42.384
20)	A Methoxychlor	6.914	7.520	319.1E6	62050439	53.710	50.063
21)	B Endrin ke...	7.147	7.663	581.0E6	105.9E6	49.032m	43.436
22)	Mirex	7.334	8.112	384.7E6	81311700	43.479	42.580

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050720\
Data File : PL058582.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 May 2020 19:16
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampled :
PSTDCCC050

Manual Integrations
APPROVED

mohammad
5/8/2020 12:59:14 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 08 00:52:59 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042820.M
Quant Title : GC Extractables
QLast Update : Wed Apr 29 00:55:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

