

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050819\
 Data File : PL048273.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 May 2019 19:27
 Operator : AJ\SJ
 Sample : K2641-02MS
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 CL-02-050719-AMS

Manual Integrations
 APPROVED

Ankita
 5/9/2019 9:24:14 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 01:04:29 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:36:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.328	3.940	213.2E6	61542701	22.014	24.707
28)	SA Decachlor...	7.996	8.976	202.6E6	58675432	20.146	22.106
Target Compounds							
2)	A alpha-BHC	3.755	4.328	985.2E6	253.8E6	63.697	72.266
3)	MA gamma-BHC...	4.030	4.612	932.5E6	232.9E6	65.626m	71.322
4)	MA Heptachlor	4.316	5.115	891.7E6	213.5E6	60.607	61.672
5)	MB Aldrin	4.555	5.418	815.4E6	190.9E6	59.513	61.657
6)	B beta-BHC	4.285	4.786	379.7E6	97086087	64.345	62.394
7)	B delta-BHC	4.479	5.000	869.9E6	194.6E6	61.614	68.798
8)	B Heptachlo...	4.992	5.804	817.6E6	192.9E6	63.457	66.151
9)	A Endosulfan I	5.322	6.160	750.7E6	179.1E6	62.400	63.353
10)	B gamma-Chl...	5.214	6.038	828.5E6	192.5E6	62.052	62.525
11)	B alpha-Chl...	5.270	6.110	800.5E6	190.5E6	61.228	62.072
12)	B 4,4'-DDE	5.438	6.270	782.7E6	178.5E6	61.704	64.645
13)	MA Dieldrin	5.561	6.416	819.6E6	190.0E6	62.946	64.775
14)	MA Endrin	5.814	6.633	732.7E6	162.2E6	60.189	62.193
15)	B Endosulfa...	6.088	6.843	635.8E6	159.9E6	55.092	59.295
16)	A 4,4'-DDD	5.949	6.760	581.8E6	146.8E6	59.769	65.405
17)	MA 4,4'-DDT	6.183	7.057	568.3E6	141.7E6	56.572	60.368
18)	B Endrin al...	6.256	6.968	511.6E6	133.9E6	52.640	57.127
19)	B Endosulfa...	6.466	7.190	590.0E6	151.3E6	54.969	58.460
20)	A Methoxychlor	6.728	7.516	290.5E6	77694309	53.351	56.911
21)	B Endrin ke...	6.949	7.658	631.5E6	162.7E6	54.186	62.270
22)	Mirex	7.123	8.109	438.0E6	117.8E6	48.865	55.879

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050819\
Data File : PL048273.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 May 2019 19:27
Operator : AJ\SJ
Sample : K2641-02MS
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
CL-02-050719-AMS

Manual Integrations
APPROVED

Ankita
5/9/2019 9:24:14 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 01:04:29 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
Quant Title : GC Extractables
QLast Update : Thu May 02 09:36:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

