

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050820\
 Data File : PL058600.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 08 May 2020 15:10
 Operator : AJ\MA
 Sample : L2506-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TR-04-050720

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 00:25:16 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042820.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 29 00:55:33 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.454	3.954	124.0E6	26955595	11.654	12.471
28)	SA Decachlor...	8.200	8.974	107.6E6	26951851	9.947	12.147

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050820\
Data File : PL058600.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 08 May 2020 15:10
Operator : AJ\MA
Sample : L2506-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
TR-04-050720

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 00:25:16 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042820.M
Quant Title : GC Extractables
QLast Update : Wed Apr 29 00:55:33 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.454 min
Delta R.T.: 0.000 min
Response: 124023066
Conc: 11.65 ng/ml

Instrument :
ECD_L
ClientSampleId :
TR-04-050720




#1 Tetrachloro-m-xylene

R.T.: 3.954 min
Delta R.T.: 0.000 min
Response: 26955595
Conc: 12.47 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.200 min
Delta R.T.: 0.003 min
Response: 107574408
Conc: 9.95 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.974 min
Delta R.T.: 0.003 min
Response: 26951851
Conc: 12.15 ng/ml