

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050919\
 Data File : PL048355.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2019 23:45
 Operator : AJ\SJ
 Sample : PB119537BS
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PLCS37

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 01:51:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050919CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri May 10 01:24:11 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.327	3.939	193.5E6	49972567	18.661	18.956
27)	SA Decachlor...	7.995	8.975	379.8E6	97860675	35.644	35.897
Target Compounds							
3)	MA gamma-BHC...	4.031	4.611	735.1E6	170.9E6	49.738	49.946
8)	B Heptachlo...	4.991	5.804	686.9E6	162.7E6	49.681	55.114
10)	B trans-Chl...	5.214	6.038	715.8E6	159.7E6	49.564	49.703
12)	B 4,4'-DDE	5.438	6.269	1385.7E6	309.3E6	98.392	100.563
13)	MA Dieldrin	5.561	6.415	1431.1E6	325.7E6	100.086	103.374
14)	MA Endrin	5.814	6.632	1318.9E6	288.5E6	105.670	109.696
19)	B Endosulfa...	6.467	7.190	993.6E6	239.1E6	86.704	90.227

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050919\
Data File : PL048355.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2019 23:45
Operator : AJ\SJ
Sample : PB119537BS
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PLCS37

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 01:51:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050919CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 10 01:24:11 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

