

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050923\
 Data File : PL082589.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2023 09:14
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 21:08:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042023.M
 Quant Title : GC Extractables
 QLast Update : Fri Apr 21 03:12:31 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.331	2.643	139.8E6	114.6E6	52.350	49.460
28)	SA Decachlor...	8.754	7.759	101.2E6	114.5E6	53.004	50.143
Target Compounds							
2)	A alpha-BHC	3.781	3.142	215.9E6	170.9E6	54.245	50.221
3)	MA gamma-BHC...	4.110	3.469	200.3E6	160.6E6	53.573	49.959
4)	MA Heptachlor	4.689	3.801	194.2E6	170.5E6	53.160	49.798
5)	MB Aldrin	5.025	4.078	178.4E6	168.2E6	54.092	53.386
6)	B beta-BHC	4.313	3.770	86876893	67305068	50.939	49.329
7)	B delta-BHC	4.555	3.996	197.4E6	156.8E6	54.866	51.114
8)	B Heptachlo...	5.451	4.581	164.1E6	144.7E6	53.783	50.220
9)	A Endosulfan I	5.833	4.948	151.5E6	135.8E6	54.657	50.857
10)	B gamma-Chl...	5.706	4.829	171.8E6	147.6E6	57.128	50.860
11)	B alpha-Chl...	5.784	4.892	163.8E6	147.1E6	54.875	50.669
12)	B 4,4'-DDE	5.963	5.083	141.6E6	122.7E6	55.688	53.373
13)	MA Dieldrin	6.107	5.212	152.4E6	138.5E6	55.219	51.298
14)	MA Endrin	6.334	5.488	127.5E6	122.3E6	52.573	50.001
15)	B Endosulfa...	6.556	5.784	123.6E6	116.5E6	50.776	51.676
16)	A 4,4'-DDD	6.478	5.638	118.3E6	103.5E6	58.341	55.335
17)	MA 4,4'-DDT	6.789	5.887	117.0E6	104.7E6	51.156	49.566
18)	B Endrin al...	6.686	5.964	101.8E6	94957880	54.037	50.742
19)	B Endosulfa...	6.920	6.187	128.1E6	120.5E6	54.009	51.170
20)	A Methoxychlor	7.270	6.466	61327343	58295528	50.255	48.105
21)	B Endrin ke...	7.400	6.692	131.7E6	133.9E6	55.001	52.742
22)	Mirex	7.864	6.863	102.5E6	115.0E6	52.380	49.894

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL050923\
Data File : PL082589.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2023 09:14
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 21:08:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL042023.M
Quant Title : GC Extractables
QLast Update : Fri Apr 21 03:12:31 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

