

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051122\
 Data File : PL074623.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 May 2022 17:04
 Operator : AR\AJ
 Sample : PB144631BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 12 01:35:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051122.M
 Quant Title : GC Extractables
 QLast Update : Wed May 11 16:19:10 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.647	4.653	11861789	27503291	23.362	20.039
28)	SA Decachlor...	11.648	10.406	12425710	24161415	27.824	23.540
Target Compounds							
2)	A alpha-BHC	6.206f	5.382f	576145	1842435	0.842	1.095 #
6)	B beta-BHC	0.000	6.129f	0	215329	N.D.	0.271 #
8)	B Heptachlo...	8.130	0.000	293314	0	0.516	N.D. #
9)	A Endosulfan I	8.509f	0.000	4968228	0	8.618	N.D. #
13)	MA Dieldrin	0.000	7.782	0	239049	N.D.	0.188 #
14)	MA Endrin	0.000	8.085f	0	217665	N.D.	0.206 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051122\
Data File : PL074623.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 May 2022 17:04
Operator : AR\AJ
Sample : PB144631BL
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 12 01:35:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051122.M
Quant Title : GC Extractables
QLast Update : Wed May 11 16:19:10 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 5.647 min
Delta R.T.: 0.000 min
Response: 11861789
Conc: 23.36 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 4.653 min
Delta R.T.: -0.002 min
Response: 27503291
Conc: 20.04 ng/ml




#2 alpha-BHC

R.T.: 6.206 min
Delta R.T.: -0.025 min
Response: 576145
Conc: 0.84 ng/ml



#2 alpha-BHC


R.T.: 5.382 min
Delta R.T.: 0.027 min
Response: 1842435
Conc: 1.09 ng/ml



#6 beta-BHC


R.T.: 0.000 min
Exp R.T.: 6.881 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#6 beta-BHC

R.T.: 6.129 min
Delta R.T.: -0.024 min
Response: 215329
Conc: 0.27 ng/ml




#8 Heptachlor epoxide

R.T.: 8.130 min
Delta R.T.: 0.003 min
Response: 293314
Conc: 0.52 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 7.086 min
Response: 0
Conc: N.D.



#9 Endosulfan I


R.T.: 8.509 min
Delta R.T.: -0.024 min
Response: 4968228
Conc: 8.62 ng/ml

Instrument :
ECD_L
ClientSampleId :




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T. : 7.488 min
Response: 0
Conc: N.D.




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 8.826 min
Response: 0
Conc: N.D.



#13 Dieldrin


R.T.: 7.782 min
Delta R.T.: 0.008 min
Response: 239049
Conc: 0.19 ng/ml



#14 Endrin


R.T.: 0.000 min
Exp R.T.: 9.067 min
Response: 0
Conc: N.D.

Instrument :
ECD_L
ClientSampleId :




#14 Endrin

R.T.: 8.085 min
Delta R.T.: 0.021 min
Response: 217665
Conc: 0.21 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.648 min
Delta R.T.: -0.001 min
Response: 12425710
Conc: 27.82 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.406 min
Delta R.T.: -0.002 min
Response: 24161415
Conc: 23.54 ng/ml