

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\
 Method File : LTX051220.M
 Title : GC Extractables
 Last Update : Wed May 13 04:29:54 2020
 Response Via : Initial Calibration

Calibration Files

500 =PL058642.D 1000 =PL058640.D 250 =PL058643.D
 750 =PL058641.D 100 =PL058644.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	1.579	1.632	1.578	1.587	1.737	1.623	E6	4.18
2)	Toxaphene-1	1.017	1.034	1.095	1.023	1.352	1.104	E4	12.86
3)	Toxaphene-2	0.956	1.018	1.028	0.978	1.046	1.005	E4	3.69
4)	Toxaphene-3	1.209	1.229	1.256	1.241	1.630	1.313	E4	13.54
5)	Toxaphene-4	2.513	2.632	2.497	2.556	2.885	2.616	E4	6.07
6)	Toxaphene-5	2.573	2.653	2.475	2.609	2.729	2.608	E4	3.62
7) SA	Decachlorobiphenyl	1.290	1.281	1.324	1.273	1.464	1.327	E6	5.97

Signal #2 Calibration Files

500 =PL058642.D 1000 =PL058640.D 250 =PL058643.D
 750 =PL058641.D 100 =PL058644.D

	Compound	500	1000	250	750	100	Avg		%RSD

1) SA	Tetrachloro-m-xylene	2.094	2.167	2.081	2.115	2.251	2.142	E6	3.25
2)	Toxaphene-1	1.108	1.137	1.136	1.145	1.240	1.153	E4	4.37
3)	Toxaphene-2	1.915	1.928	1.834	1.879	2.105	1.932	E4	5.34
4)	Toxaphene-3	5.244	5.152	5.036	5.100	5.585	5.223	E4	4.13
5)	Toxaphene-4	3.123	3.084	3.018	3.003	3.309	3.107	E4	3.95
6)	Toxaphene-5	2.556	2.540	2.546	2.471	2.666	2.556	E4	2.75
7) SA	Decachlorobiphenyl	1.717	1.668	1.782	1.667	2.030	1.773	E6	8.54

(#) = Out of Range