

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051220\
 Data File : PL058630.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 May 2020 21:03
 Operator : AJ\MA
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 02:26:26 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051220.M
 Quant Title : GC Extractables
 QLast Update : Wed May 13 02:26:09 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.533	4.078	160.5E6	214.5E6	98.453	99.291
28)	SA Decachlor...	8.324	9.193	123.6E6	163.5E6	95.869	94.664
Target Compounds							
2)	A alpha-BHC	3.976	4.471	240.2E6	324.1E6	103.139	103.700
3)	MA gamma-BHC...	4.263	4.759	213.0E6	303.8E6	102.412	102.059
4)	MA Heptachlor	4.572	5.273	189.5E6	281.2E6	100.810	100.994
5)	MB Aldrin	4.823	5.581	170.5E6	254.7E6	101.801	102.148
6)	B beta-BHC	4.512	4.926	83879345	126.7E6	96.466	96.693
7)	B delta-BHC	4.720	5.146	196.2E6	293.9E6	103.775	102.904
8)	B Heptachlo...	5.268	5.969	156.1E6	236.2E6	99.534	99.656
9)	A Endosulfan I	5.610	6.328	143.2E6	211.9E6	99.292	99.557
10)	B gamma-Chl...	5.497	6.203	155.3E6	231.6E6	100.316	99.853
11)	B alpha-Chl...	5.556	6.277	153.9E6	229.5E6	99.652	98.907
12)	B 4,4'-DDE	5.720	6.431	144.6E6	214.3E6	103.103	102.138
13)	MA Dieldrin	5.855	6.586	151.9E6	221.7E6	101.988	101.612
14)	MA Endrin	6.113	6.805	133.0E6	185.7E6	102.584	102.700
15)	B Endosulfa...	6.385	7.013	136.1E6	191.7E6	99.723	98.814
16)	A 4,4'-DDD	6.237	6.923	121.1E6	176.3E6	101.796	101.049
17)	MA 4,4'-DDT	6.480	7.223	125.8E6	173.4E6	103.211	102.074
18)	B Endrin al...	6.554	7.138	113.1E6	164.5E6	98.144	97.651
19)	B Endosulfa...	6.768	7.361	131.1E6	182.2E6	99.572	98.586
20)	A Methoxychlor	7.028	7.679	64077155	95690436	98.375	97.840
21)	B Endrin ke...	7.260	7.834	148.4E6	198.0E6	99.120	98.809
22)	Mirex	7.457	8.292	113.4E6	143.4E6	94.092	94.229

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051220\
Data File : PL058630.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 May 2020 21:03
Operator : AJ\MA
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 02:26:26 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051220.M
Quant Title : GC Extractables
QLast Update : Wed May 13 02:26:09 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

