

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051325\
 Data File : PL095630.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 May 2025 11:43
 Operator : AR\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 13 14:41:49 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051325.M
 Quant Title : GC Extractables
 QLast Update : Tue May 13 14:38:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.594	2.911	65069659	75681262	15.925	16.157
28)	SA Decachlor...	9.118	8.088	64678160	75992234	19.065	18.424
Target Compounds							
8)	B Heptachlo...	5.726	0.000	12636820	0	2.439	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051325\
Data File : PL095630.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 May 2025 11:43
Operator : AR\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 13 14:41:49 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051325.M
Quant Title : GC Extractables
QLast Update : Tue May 13 14:38:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.594 min
Delta R.T.: 0.001 min
Response: 65069659
Conc: 15.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 2.911 min
Delta R.T.: 0.001 min
Response: 75681262
Conc: 16.16 ng/ml




#8 Heptachlor epoxide

R.T.: 5.726 min
Delta R.T.: -0.009 min
Response: 12636820
Conc: 2.44 ng/ml



#8 Heptachlor epoxide


R.T.: 0.000 min
Exp R.T.: 4.895 min
Response: 0
Conc: N.D.



#28 Decachlorobiphenyl

R.T.: 9.118 min
Delta R.T.: 0.002 min
Response: 64678160
Conc: 19.07 ng/ml

Instrument :
ECD_L
ClientSampleId :
I.BLK



#28 Decachlorobiphenyl

R.T.: 8.088 min
Delta R.T.: 0.001 min
Response: 75992234
Conc: 18.42 ng/ml