

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051719\
 Data File : PL048605.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 May 2019 19:56
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 01:07:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:36:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.327	3.940	459.4E6	123.6E6	47.425	49.609
2)	SA Decachlor...	7.995	8.978	464.5E6	126.7E6	46.199	47.717
Target Compounds							
2)	A alpha-BHC	3.755	4.328	756.6E6	179.5E6	48.917	51.110
3)	MA gamma-BHC...	4.031	4.611	693.5E6	164.6E6	48.805	50.380
4)	MA Heptachlor	4.316	5.115	697.2E6	165.9E6	47.388	47.927
5)	MB Aldrin	4.555	5.418	663.2E6	152.9E6	48.405	49.388
6)	B beta-BHC	4.285	4.786	283.5E6	75799220	48.043	48.714
7)	B delta-BHC	4.478	5.000	695.3E6	150.7E6	49.249	53.285
8)	B Heptachlo...	4.991	5.804	602.7E6	143.7E6	46.776	49.262
9)	A Endosulfan I	5.321	6.160	563.5E6	133.9E6	46.845	47.369
10)	B gamma-Chl...	5.213	6.038	625.5E6	147.4E6	46.848	47.866
11)	B alpha-Chl...	5.270	6.111	608.3E6	145.8E6	46.531	47.534
12)	B 4,4'-DDE	5.437	6.270	542.1E6	133.7E6	42.732	48.430
13)	MA Dieldrin	5.561	6.417	606.4E6	140.4E6	46.570	47.856
14)	MA Endrin	5.813	6.633	519.3E6	112.9E6	42.654	43.284
15)	B Endosulfa...	6.088	6.845	521.7E6	125.4E6	45.210	46.490
16)	A 4,4'-DDD	5.948	6.761	476.5E6	113.5E6	48.949	50.567
17)	MA 4,4'-DDT	6.182	7.058	404.9E6	98083232	40.298	41.798
18)	B Endrin al...	6.255	6.969	435.1E6	108.6E6	44.773	46.326
19)	B Endosulfa...	6.467	7.191	477.7E6	119.5E6	44.505	46.165
20)	A Methoxychlor	6.728	7.518	209.6E6	57093156	38.488	41.821
21)	B Endrin ke...	6.949	7.660	543.5E6	129.6E6	46.633	49.597
22)	Mirex	7.123	8.110	376.3E6	98505392	41.977	46.710

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051719\
Data File : PL048605.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 May 2019 19:56
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 01:07:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
Quant Title : GC Extractables
QLast Update : Thu May 02 09:36:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

