

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051819\
 Data File : PL048642.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 May 2019 12:17
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 17 23:14:10 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:36:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.327	3.939	487.9E6	130.6E6	50.367	52.426
2)	SA Decachlor...	7.997	8.977	494.7E6	134.8E6	49.207	50.792
Target Compounds							
2)	A alpha-BHC	3.755	4.328	807.5E6	190.1E6	52.208	54.115
3)	MA gamma-BHC...	4.031	4.611	740.6E6	174.3E6	52.125	53.368
4)	MA Heptachlor	4.316	5.115	743.0E6	175.4E6	50.501	50.674
5)	MB Aldrin	4.556	5.418	711.7E6	162.6E6	51.940	52.524
6)	B beta-BHC	4.285	4.786	302.8E6	79864239	51.308	51.326
7)	B delta-BHC	4.479	5.000	742.9E6	165.8E6	52.619	58.628
8)	B Heptachlo...	4.992	5.805	644.3E6	154.1E6	50.006	52.849
9)	A Endosulfan I	5.323	6.160	601.0E6	142.3E6	49.958	50.311
10)	B gamma-Chl...	5.215	6.039	670.1E6	156.3E6	50.189	50.756
11)	B alpha-Chl...	5.270	6.110	650.6E6	154.7E6	49.762	50.425
12)	B 4,4'-DDE	5.439	6.270	577.6E6	142.1E6	45.533	51.470
13)	MA Dieldrin	5.562	6.416	658.6E6	149.4E6	50.583	50.915
14)	MA Endrin	5.814	6.633	557.7E6	120.3E6	45.810	46.122
15)	B Endosulfa...	6.090	6.845	556.7E6	132.8E6	48.243	49.248
16)	A 4,4'-DDD	5.949	6.760	515.7E6	121.0E6	52.978	53.940
17)	MA 4,4'-DDT	6.184	7.058	427.6E6	102.9E6	42.567	43.860
18)	B Endrin al...	6.256	6.969	462.3E6	114.4E6	47.568	48.827
19)	B Endosulfa...	6.468	7.191	507.5E6	126.3E6	47.287	48.783
20)	A Methoxychlor	6.729	7.518	276.3E6	74238622	50.744	54.380
21)	B Endrin ke...	6.950	7.660	581.8E6	137.4E6	49.921	52.597
22)	Mirex	7.124	8.110	399.2E6	104.2E6	44.536	49.432



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051819\
Data File : PL048642.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 May 2019 12:17
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 17 23:14:10 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
Quant Title : GC Extractables
QLast Update : Thu May 02 09:36:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.327 min
Delta R.T.: 0.000 min
Response: 487896973
Conc: 50.37 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#1 Tetrachloro-m-xylene

R.T.: 3.939 min
Delta R.T.: 0.000 min
Response: 130586884
Conc: 52.43 ng/ml




#2 alpha-BHC

R.T.: 3.755 min
Delta R.T.: 0.000 min
Response: 807527743
Conc: 52.21 ng/ml



#2 alpha-BHC


R.T.: 4.328 min
Delta R.T.: 0.000 min
Response: 190050824
Conc: 54.11 ng/ml



#3 gamma-BHC (Lindane)


R.T.: 4.031 min
Delta R.T.: 0.000 min
Response: 740643903
Conc: 52.12 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#3 gamma-BHC (Lindane)

R.T.: 4.611 min
Delta R.T.: 0.000 min
Response: 174309000
Conc: 53.37 ng/ml




#4 Heptachlor

R.T.: 4.316 min
Delta R.T.: 0.000 min
Response: 742968705
Conc: 50.50 ng/ml



#4 Heptachlor


R.T.: 5.115 min
Delta R.T.: 0.000 min
Response: 175423281
Conc: 50.67 ng/ml



#5 Aldrin


R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 711660441
Conc: 51.94 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#5 Aldrin

R.T.: 5.418 min
Delta R.T.: 0.000 min
Response: 162648709
Conc: 52.52 ng/ml




#6 beta-BHC

R.T.: 4.285 min
Delta R.T.: 0.000 min
Response: 302796299
Conc: 51.31 ng/ml



#6 beta-BHC


R.T.: 4.786 min
Delta R.T.: 0.000 min
Response: 79864239
Conc: 51.33 ng/ml



#7 delta-BHC


R.T.: 4.479 min
Delta R.T.: 0.000 min
Response: 742916764
Conc: 52.62 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#7 delta-BHC

R.T.: 5.000 min
Delta R.T.: 0.000 min
Response: 165794035
Conc: 58.63 ng/ml




#8 Heptachlor epoxide

R.T.: 4.992 min
Delta R.T.: 0.000 min
Response: 644310752
Conc: 50.01 ng/ml



#8 Heptachlor epoxide


R.T.: 5.805 min
Delta R.T.: 0.000 min
Response: 154146840
Conc: 52.85 ng/ml



#9 Endosulfan I


R.T.: 5.323 min
Delta R.T.: 0.000 min
Response: 600987680
Conc: 49.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#9 Endosulfan I

R.T.: 6.160 min
Delta R.T.: 0.000 min
Response: 142253291
Conc: 50.31 ng/ml




#10 gamma-Chlordane

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 670133653
Conc: 50.19 ng/ml



#10 gamma-Chlordane


R.T.: 6.039 min
Delta R.T.: 0.000 min
Response: 156259650
Conc: 50.76 ng/ml



#11 alpha-Chlordane


R.T.: 5.270 min
Delta R.T.: 0.000 min
Response: 650560667
Conc: 49.76 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#11 alpha-Chlordane

R.T.: 6.110 min
Delta R.T.: 0.000 min
Response: 154719362
Conc: 50.43 ng/ml




#12 4,4'-DDE

R.T.: 5.439 min
Delta R.T.: 0.000 min
Response: 577597842
Conc: 45.53 ng/ml



#12 4,4'-DDE


R.T.: 6.270 min
Delta R.T.: 0.000 min
Response: 142103817
Conc: 51.47 ng/ml



#13 Dieldrin


R.T.: 5.562 min
Delta R.T.: 0.000 min
Response: 658623566
Conc: 50.58 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#13 Dieldrin

R.T.: 6.416 min
Delta R.T.: 0.000 min
Response: 149368502
Conc: 50.92 ng/ml




#14 Endrin

R.T.: 5.814 min
Delta R.T.: 0.000 min
Response: 557688331
Conc: 45.81 ng/ml



#14 Endrin


R.T.: 6.633 min
Delta R.T.: 0.000 min
Response: 120259825
Conc: 46.12 ng/ml



#15 Endosulfan II


R.T.: 6.090 min
Delta R.T.: 0.000 min
Response: 556747759
Conc: 48.24 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#15 Endosulfan II

R.T.: 6.845 min
Delta R.T.: 0.002 min
Response: 132802815
Conc: 49.25 ng/ml




#16 4,4'-DDD

R.T.: 5.949 min
Delta R.T.: 0.000 min
Response: 515692176
Conc: 52.98 ng/ml



#16 4,4'-DDD


R.T.: 6.760 min
Delta R.T.: 0.001 min
Response: 121043643
Conc: 53.94 ng/ml



#17 4,4'-DDT


R.T.: 6.184 min
Delta R.T.: 0.000 min
Response: 427649003
Conc: 42.57 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#17 4,4'-DDT

R.T.: 7.058 min
Delta R.T.: 0.001 min
Response: 102922998
Conc: 43.86 ng/ml




#18 Endrin aldehyde

R.T.: 6.256 min
Delta R.T.: 0.000 min
Response: 462280977
Conc: 47.57 ng/ml



#18 Endrin aldehyde


R.T.: 6.969 min
Delta R.T.: 0.001 min
Response: 114423533
Conc: 48.83 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.468 min
Delta R.T.: 0.000 min
Response: 507512051
Conc: 47.29 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#19 Endosulfan Sulfate

R.T.: 7.191 min
Delta R.T.: 0.001 min
Response: 126260838
Conc: 48.78 ng/ml




#20 Methoxychlor

R.T.: 6.729 min
Delta R.T.: 0.000 min
Response: 276289392
Conc: 50.74 ng/ml



#20 Methoxychlor


R.T.: 7.518 min
Delta R.T.: 0.002 min
Response: 74238622
Conc: 54.38 ng/ml



#21 Endrin ketone


R.T.: 6.950 min
Delta R.T.: 0.000 min
Response: 581830030
Conc: 49.92 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050




#21 Endrin ketone

R.T.: 7.660 min
Delta R.T.: 0.002 min
Response: 137425258
Conc: 52.60 ng/ml




#22 Mirex

R.T.: 7.124 min
Delta R.T.: 0.000 min
Response: 399213211
Conc: 44.54 ng/ml



#22 Mirex


R.T.: 8.110 min
Delta R.T.: 0.001 min
Response: 104245987
Conc: 49.43 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.997 min
Delta R.T.: 0.000 min
Response: 494734738
Conc: 49.21 ng/ml

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050



#28 Decachlorobiphenyl

R.T.: 8.977 min
Delta R.T.: 0.002 min
Response: 134817546
Conc: 50.79 ng/ml