

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051923\
 Data File : PL082809.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2023 13:44
 Operator : AR\AJ
 Sample : PSTDICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 04:13:08 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:12:52 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.317	2.639	150.9E6	135.1E6	50.000	50.000
28)	SA Decachlor...	8.733	7.744	103.2E6	122.3E6	50.000	50.000
Target Compounds							
2)	A alpha-BHC	3.766	3.135	235.9E6	198.1E6	50.000	50.000
3)	MA gamma-BHC...	4.094	3.462	215.3E6	173.8E6	50.000	50.000
4)	MA Heptachlor	4.672	3.793	196.7E6	190.6E6	50.000	50.000
5)	MB Aldrin	5.008	4.070	184.6E6	170.0E6	50.000	50.000
6)	B beta-BHC	4.298	3.762	92837968	73581063	50.000	50.000
7)	B delta-BHC	4.538	3.988	177.6E6	173.8E6	50.000	50.000
8)	B Heptachlo...	5.434	4.570	159.9E6	153.1E6	50.000	50.000
9)	A Endosulfan I	5.815	4.937	150.7E6	144.1E6	50.000	50.000
10)	B gamma-Chl...	5.689	4.819	161.6E6	154.3E6	50.000	50.000
11)	B alpha-Chl...	5.766	4.882	160.8E6	154.5E6	50.000	50.000
12)	B 4,4'-DDE	5.945	5.072	141.6E6	127.0E6	50.000	50.000
13)	MA Dieldrin	6.089	5.201	153.9E6	146.7E6	50.000	50.000
14)	MA Endrin	6.316	5.475	136.8E6	131.1E6	50.000	50.000
15)	B Endosulfa...	6.539	5.772	135.6E6	123.7E6	50.000	50.000
16)	A 4,4'-DDD	6.461	5.626	112.0E6	100.7E6	50.000	50.000
17)	MA 4,4'-DDT	6.772	5.875	127.9E6	116.2E6	50.000	50.000
18)	B Endrin al...	6.668	5.952	101.7E6	99506413	50.000	50.000
19)	B Endosulfa...	6.903	6.174	129.1E6	124.9E6	50.000	50.000
20)	A Methoxychlor	7.252	6.454	66712514	64847775	50.000	50.000
21)	B Endrin ke...	7.382	6.678	130.5E6	138.6E6	50.000	50.000
22)	Mirex	7.846	6.849	103.8E6	118.1E6	50.000	50.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051923\
Data File : PL082809.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2023 13:44
Operator : AR\AJ
Sample : PSTDICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 04:13:08 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:12:52 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

