

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051923\
 Data File : PL082836.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 May 2023 20:54
 Operator : AR\AJ
 Sample : 02848-01
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP4

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/22/2023
 Supervised By : Ankita Jodhani 05/22/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 05:04:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.316	2.637	41040093	32673610	13.964m	12.586m
28) SA Decachlor...	8.730	7.743	24160176	30490092	11.520	12.262

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL051923\
Data File : PL082836.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 May 2023 20:54
Operator : AR\AJ
Sample : 02848-01
Misc :
ALS Vial : 31 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
TP4


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/22/2023
Supervised By : Ankita Jodhani 05/22/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 05:04:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 3.316 min
Delta R.T.: -0.001 min
Response: 41040093
Conc: 13.96 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP4


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/22/2023
Supervised By :Ankita Jodhani 05/22/2023




#1 Tetrachloro-m-xylene

R.T.: 2.637 min
Delta R.T.: -0.002 min
Response: 32673610
Conc: 12.59 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.730 min
Delta R.T.: -0.003 min
Response: 24160176
Conc: 11.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 7.743 min
Delta R.T.: 0.000 min
Response: 30490092
Conc: 12.26 ng/ml