

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052020\
 Data File : PL058809.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 May 2020 11:50
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 20 17:24:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052020.M
 Quant Title : GC Extractables
 QLast Update : Wed May 20 11:27:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.530	4.077	88188511	134.1E6	48.095	49.517
2)	SA Decachlor...	8.312	9.188	80659019	126.6E6	47.739	48.024
Target Compounds							
2)	A alpha-BHC	3.972	4.469	129.3E6	203.4E6	47.575	50.672
3)	MA gamma-BHC...	4.258	4.756	115.0E6	191.1E6	48.733	50.418
4)	MA Heptachlor	4.565	5.270	106.8E6	189.0E6	47.520	49.330
5)	MB Aldrin	4.817	5.578	93075314	170.8E6	48.076	50.066
6)	B beta-BHC	4.507	4.924	48207471	81801777	46.906	47.011
7)	B delta-BHC	4.715	5.144	105.1E6	182.9E6	49.005	50.144
8)	B Heptachlo...	5.262	5.966	88272043	162.0E6	47.827	49.129
9)	A Endosulfan I	5.603	6.325	81364929	141.8E6	47.932	48.249
10)	B gamma-Chl...	5.490	6.200	87806639	160.0E6	47.894	48.309
11)	B alpha-Chl...	5.549	6.274	87597407	157.0E6	47.873	48.338
12)	B 4,4'-DDE	5.713	6.428	79045175	134.5E6	48.137	47.115
13)	MA Dieldrin	5.847	6.583	84451307	146.5E6	48.111	48.359
14)	MA Endrin	6.105	6.802	73674639	121.4E6	47.008	48.632
15)	B Endosulfa...	6.378	7.009	79047817	132.6E6	47.446	47.294
16)	A 4,4'-DDD	6.229	6.919	68494287	119.4E6	48.388	50.051
17)	MA 4,4'-DDT	6.472	7.219	69707022	117.9E6	47.921	49.571
18)	B Endrin al...	6.545	7.133	67299113	115.4E6	47.689	48.920
19)	B Endosulfa...	6.759	7.356	77364713	128.5E6	47.420	48.898
20)	A Methoxychlor	7.019	7.676	40979387	68454122	48.040	49.083
21)	B Endrin ke...	7.251	7.830	90394217	137.6E6	49.017	49.679
22)	Mirex	7.446	8.288	72471098	104.9E6	46.857	47.855

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052020\
Data File : PL058809.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 May 2020 11:50
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 20 17:24:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052020.M
Quant Title : GC Extractables
QLast Update : Wed May 20 11:27:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

