

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052219\
 Data File : PL048850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 May 2019 18:09
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 23 01:12:22 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
 Quant Title : GC Extractables
 QLast Update : Thu May 02 09:36:55 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	3.937	469.6E6	128.2E6	48.478	51.472
28)	SA Decachlor...	7.992	8.972	533.8E6	142.6E6	53.093	53.728
Target Compounds							
2)	A alpha-BHC	3.753	4.325	775.8E6	187.7E6	50.154	53.448
3)	MA gamma-BHC...	4.029	4.608	724.3E6	171.9E6	50.975	52.629
4)	MA Heptachlor	4.314	5.112	733.9E6	176.9E6	49.885	51.105
5)	MB Aldrin	4.553	5.414	698.4E6	160.8E6	50.969	51.925
6)	B beta-BHC	4.283	4.783	290.6E6	76004570	49.237	48.846
7)	B delta-BHC	4.477	4.997	726.4E6	139.5E6	51.450	49.345
8)	B Heptachlo...	4.989	5.800	633.0E6	136.0E6	49.131	46.615
9)	A Endosulfan I	5.319	6.156	593.9E6	142.2E6	49.372	50.280
10)	B gamma-Chl...	5.211	6.034	659.7E6	154.1E6	49.410	50.053
11)	B alpha-Chl...	5.267	6.106	635.9E6	154.2E6	48.639	50.242
12)	B 4,4'-DDE	5.435	6.266	595.7E6	144.0E6	46.960	52.170
13)	MA Dieldrin	5.559	6.412	657.9E6	150.1E6	50.530	51.159
14)	MA Endrin	5.811	6.629	577.0E6	123.9E6	47.393	47.501
15)	B Endosulfa...	6.085	6.841	561.0E6	136.6E6	48.608	50.639
16)	A 4,4'-DDD	5.946	6.756	490.1E6	118.1E6	50.344	52.630
17)	MA 4,4'-DDT	6.180	7.053	463.9E6	113.6E6	46.175	48.391
18)	B Endrin al...	6.253	6.964	466.5E6	115.8E6	48.001	49.408
19)	B Endosulfa...	6.464	7.187	516.7E6	127.1E6	48.145	49.104
20)	A Methoxychlor	6.725	7.514	243.5E6	65258567	44.719	47.802
21)	B Endrin ke...	6.946	7.655	586.9E6	137.3E6	50.356	52.568
22)	Mirex	7.119	8.105	403.3E6	109.0E6	44.994	51.683

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052219\
Data File : PL048850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 May 2019 18:09
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 23 01:12:22 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL050119.M
Quant Title : GC Extractables
QLast Update : Thu May 02 09:36:55 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

