

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052423\
 Data File : PL082969.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2023 19:04
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 25 05:29:57 2023
 Quant Title :
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
1) SA Tetrachloro-m-xylene	3.320	135878526	46.233 ng/ml
28) SA Decachlorobiphenyl	8.737	94289280	44.957 ng/ml
30) SA Tetrachloro-m-xylene #2	2.636	118060921	45.476 ng/ml
57) SA Decachlorobiphenyl #2	7.743	108936514	43.810 ng/ml
Target Compounds			
2) A alpha-BHC	3.769	208621526	46.995 ng/ml
3) MA gamma-BHC (Lindane)	4.097	192576604	46.483 ng/ml
4) MA Heptachlor	4.674	177885319	46.537 ng/ml
5) MB Aldrin	5.011	169883982	48.288 ng/ml
6) B beta-BHC	4.301	82538988	43.840 ng/ml
7) B delta-BHC	4.542	179663939	52.244 ng/ml
8) B Heptachlor epoxide	5.437	156986133	49.078 ng/ml
9) A Endosulfan I	5.817	163619567	54.529 ng/ml
10) B gamma-Chlordane	5.690	156529500	49.270 ng/mlm
11) B alpha-Chlordane	5.768	156243740	48.685 ng/mlm
12) B 4,4'-DDE	5.948	133000391	48.250 ng/ml
13) MA Dieldrin	6.092	143977211	47.867 ng/ml
14) MA Endrin	6.319	126709086	47.152 ng/ml
15) B Endosulfan II	6.542	131361250	49.645 ng/ml
16) A 4,4'-DDD	6.464	107514901	49.476 ng/ml
17) MA 4,4'-DDT	6.775	117710212	47.024 ng/ml
18) B Endrin aldehyde	6.672	95290475	47.393 ng/ml
19) B Endosulfan Sulfate	6.906	121183795	47.269 ng/ml
20) A Methoxychlor	7.256	61413452	46.442 ng/ml
21) B Endrin ketone	7.385	122736755	48.117 ng/ml
22) Mirex	7.849	96948093	46.253 ng/ml
31) A alpha-BHC #2	3.132	172009846	46.097 ng/ml
32) MA gamma-BHC (Lindane) #2	3.459	156334901	46.901 ng/ml
33) MA Heptachlor #2	3.789	150186607	41.022 ng/ml
34) MB Aldrin #2	4.066	158277190	48.646 ng/ml
35) B beta-BHC #2	3.759	65066297	45.498 ng/ml
36) B delta-BHC #2	3.985	155492813	47.458 ng/ml
37) B Heptachlor epoxide #2	4.567	136181122	45.885 ng/ml
38) A Endosulfan I #2	4.934	122161923	43.717 ng/ml
39) B gamma-Chlordane #2	4.816	138015134	45.953 ng/ml
40) B alpha-Chlordane #2	4.878	138022510	45.977 ng/ml
41) B 4,4'-DDE #2	5.068	114857412	47.291 ng/mlm
42) MA Dieldrin #2	5.198	130670637	46.289 ng/ml
43) MA Endrin #2	5.473	118144562	46.569 ng/ml
44) B Endosulfan II #2	5.769	108204685	43.733 ng/ml
45) A 4,4'-DDD #2	5.623	89935822	46.763 ng/ml
46) MA 4,4'-DDT #2	5.872	99596175	44.641 ng/ml
47) B Endrin aldehyde #2	5.949	88018464	44.511 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052423\
Data File : PL082969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2023 19:04
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 05:29:57 2023
Quant Title :
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	R.T.	Response	Conc Units
48) B	Endosulfan Sulfate #2	6.172	110639594	45.178 ng/ml
49) A	Methoxychlor #2	6.451	56394988	44.344 ng/ml
50) B	Endrin ketone #2	6.675	112995871	41.791 ng/mlm
51)	Mirex #2	6.847	104972793	44.863 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052423\
Data File : PL082969.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2023 19:04
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 25 05:29:57 2023
Quant Title :
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

