

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052424\
 Data File : PL089709.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2024 11:01
 Operator : AR\AJ
 Sample : PSTDICC100
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDICC100

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 14:41:22 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052424.M
 Quant Title : GC Extractables
 QLast Update : Fri May 24 14:38:12 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.549	2.794	203.2E6	149.5E6	99.932	102.052
2)	SA Decachlor...	9.060	7.941	136.6E6	146.8E6	96.990	99.810
Target Compounds							
2)	A alpha-BHC	4.004	3.298	299.0E6	239.7E6	101.732	104.614
3)	MA gamma-BHC...	4.336	3.629	291.4E6	229.7E6	101.484	102.693
4)	MA Heptachlor	4.925	3.970	249.5E6	215.4E6	99.966	103.228
5)	MB Aldrin	5.267	4.251	254.8E6	220.3E6	100.268	104.091
6)	B beta-BHC	4.531	3.926	122.9E6	93165025	98.693	101.162
7)	B delta-BHC	4.778	4.157	272.1E6	226.2E6	101.288	104.689
8)	B Heptachlo...	5.692	4.754	230.9E6	193.9E6	99.143	102.461
9)	A Endosulfan I	6.078	5.124	210.8E6	180.2E6	98.756	102.484
10)	B gamma-Chl...	5.948	5.004	226.0E6	200.0E6	99.393	103.405
11)	B alpha-Chl...	6.027	5.068	224.7E6	199.1E6	99.073	102.850
12)	B 4,4'-DDE	6.199	5.257	205.4E6	185.5E6	99.700	104.473
13)	MA Dieldrin	6.353	5.388	225.3E6	199.3E6	99.630	103.586
14)	MA Endrin	6.582	5.664	192.3E6	176.7E6	99.565	103.435
15)	B Endosulfa...	6.800	5.957	194.9E6	167.4E6	98.526	102.091
16)	A 4,4'-DDD	6.715	5.811	166.4E6	152.4E6	99.043	103.647
17)	MA 4,4'-DDT	7.029	6.062	174.7E6	166.2E6	98.818	103.903
18)	B Endrin al...	6.930	6.136	145.1E6	125.7E6	98.392	101.925
19)	B Endosulfa...	7.164	6.359	175.7E6	161.6E6	97.584	101.573
20)	A Methoxychlor	7.504	6.637	92480648	89387324	97.263	99.968
21)	B Endrin ke...	7.648	6.865	198.3E6	190.2E6	98.421	101.493
22)	Mirex	8.123	7.048	151.2E6	149.7E6	96.796	99.723

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052424\
Data File : PL089709.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2024 11:01
Operator : AR\AJ
Sample : PSTDICC100
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDICC100

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 14:41:22 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052424.M
Quant Title : GC Extractables
QLast Update : Fri May 24 14:38:12 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

