

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052424\
 Data File : PL089708.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 May 2024 10:47
 Operator : AR\AJ
 Sample : RESCHK
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 RESCHK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 24 15:00:39 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052424.M
 Quant Title : GC Extractables
 QLast Update : Fri May 24 14:57:01 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.548	2.794	41737598	27236796	19.934	18.996
28)	SA Decachlor...	9.060	7.940	29736587	29058048	20.224	19.554
Target Compounds							
9)	A Endosulfan I	6.077	5.124	23083412	14713572	10.221	8.613
10)	B gamma-Chl...	5.947	5.004	26485636	17169396	10.952	9.212
12)	B 4,4'-DDE	6.199	5.257	42530542	30217766	19.551	18.087
13)	MA Dieldrin	6.352	5.388	46464891	32843797	19.600	17.824
19)	B Endosulfa...	7.164	6.359	38252139	29485637	19.988	18.749
20)	A Methoxychlor	7.504	6.637	84178800	79978860	85.706	89.906
21)	B Endrin ke...	7.648	6.865	40747202	33060691	19.351	18.142



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052424\
Data File : PL089708.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 May 2024 10:47
Operator : AR\AJ
Sample : RESCHK
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
RESCHK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 24 15:00:39 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052424.M
Quant Title : GC Extractables
QLast Update : Fri May 24 14:57:01 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.548 min
Delta R.T.: -0.002 min
Response: 41737598
Conc: 19.93 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#1 Tetrachloro-m-xylene

R.T.: 2.794 min
Delta R.T.: 0.000 min
Response: 27236796
Conc: 19.00 ng/ml




#9 Endosulfan I

R.T.: 6.077 min
Delta R.T.: -0.004 min
Response: 23083412
Conc: 10.22 ng/ml



#9 Endosulfan I


R.T.: 5.124 min
Delta R.T.: -0.002 min
Response: 14713572
Conc: 8.61 ng/ml



#10 gamma-Chlordane


R.T.: 5.947 min
Delta R.T.: -0.003 min
Response: 26485636
Conc: 10.95 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#10 gamma-Chlordane

R.T.: 5.004 min
Delta R.T.: -0.002 min
Response: 17169396
Conc: 9.21 ng/ml




#12 4,4'-DDE

R.T.: 6.199 min
Delta R.T.: -0.003 min
Response: 42530542
Conc: 19.55 ng/ml



#12 4,4'-DDE


R.T.: 5.257 min
Delta R.T.: -0.002 min
Response: 30217766
Conc: 18.09 ng/ml



#13 Dieldrin


R.T.: 6.352 min
Delta R.T.: -0.004 min
Response: 46464891
Conc: 19.60 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#13 Dieldrin

R.T.: 5.388 min
Delta R.T.: -0.002 min
Response: 32843797
Conc: 17.82 ng/ml




#19 Endosulfan Sulfate

R.T.: 7.164 min
Delta R.T.: -0.003 min
Response: 38252139
Conc: 19.99 ng/ml



#19 Endosulfan Sulfate


R.T.: 6.359 min
Delta R.T.: -0.003 min
Response: 29485637
Conc: 18.75 ng/ml



#20 Methoxychlor


R.T.: 7.504 min
Delta R.T.: -0.003 min
Response: 84178800
Conc: 85.71 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK




#20 Methoxychlor

R.T.: 6.637 min
Delta R.T.: -0.003 min
Response: 79978860
Conc: 89.91 ng/ml




#21 Endrin ketone

R.T.: 7.648 min
Delta R.T.: -0.003 min
Response: 40747202
Conc: 19.35 ng/ml



#21 Endrin ketone


R.T.: 6.865 min
Delta R.T.: -0.004 min
Response: 33060691
Conc: 18.14 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.060 min
Delta R.T.: -0.002 min
Response: 29736587
Conc: 20.22 ng/ml

Instrument :
ECD_L
ClientSampleId :
RESCHK



#28 Decachlorobiphenyl

R.T.: 7.940 min
Delta R.T.: -0.004 min
Response: 29058048
Conc: 19.55 ng/ml