

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052623\
 Data File : PL083065.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 26 May 2023 20:39
 Operator : AR\AJ
 Sample : PB153077BS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB153077BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 26 22:02:02 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.316	2.636	68717988	61976058	23.381	23.873
2)	SA Decachlor...	8.729	7.738	55121898	67903746	26.282	27.308
Target Compounds							
2)	A alpha-BHC	3.765	3.132	231.4E6	197.7E6	52.120	52.986
3)	MA gamma-BHC...	4.093	3.458	215.2E6	182.9E6	51.947	54.858
4)	MA Heptachlor	4.670	3.789	201.3E6	189.5E6	52.671	51.754
5)	MB Aldrin	5.006	4.064	190.4E6	173.8E6	54.123	53.405
6)	B beta-BHC	4.297	3.759	94723516	75722648	50.312	52.950
7)	B delta-BHC	4.537	3.984	204.7E6	177.7E6	59.525	54.225
8)	B Heptachlo...	5.432	4.565	171.1E6	160.5E6	53.482	54.063
9)	A Endosulfan I	5.813	4.932	158.7E6	150.8E6	52.900	53.968
10)	B gamma-Chl...	5.687	4.814	175.1E6	162.0E6	55.119	53.952
11)	B alpha-Chl...	5.765	4.877	170.8E6	161.8E6	53.205	53.895
12)	B 4,4'-DDE	5.943	5.067	151.1E6	138.0E6	54.813	56.814
13)	MA Dieldrin	6.087	5.196	162.2E6	152.6E6	53.916	54.048
14)	MA Endrin	6.314	5.470	142.5E6	134.4E6	53.024	52.967
15)	B Endosulfa...	6.537	5.767	134.1E6	128.5E6	50.786	51.933
16)	A 4,4'-DDD	6.460	5.621	124.5E6	111.6E6	57.304	58.043
17)	MA 4,4'-DDT	6.769	5.869	128.7E6	116.3E6	51.428	52.143
18)	B Endrin al...	6.667	5.946	104.9E6	102.6E6	52.154	51.909
19)	B Endosulfa...	6.901	6.169	136.7E6	131.6E6	53.336	53.743
20)	A Methoxychlor	7.250	6.448	68357911	66035671	51.693	51.925
21)	B Endrin ke...	7.380	6.673	140.1E6	146.5E6	54.928	54.174
22)	Mirex	7.844	6.843	115.4E6	129.0E6	55.040	55.148

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052623\
Data File : PL083065.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 26 May 2023 20:39
Operator : AR\AJ
Sample : PB153077BS
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PB153077BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 26 22:02:02 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

