

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052722\
 Data File : PL075126.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 May 2022 17:27
 Operator : AR\AJ
 Sample : PB145085BL
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PB145085BL

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 05/31/2022
 Supervised By : Ankita Jodhani 05/31/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 02:47:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051122.M
 Quant Title : GC Extractables
 QLast Update : Sat May 14 02:37:15 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.644	4.644	11491107	31915646	22.632	23.253
28)	SA Decachlor...	11.639	10.390	10539419	24491419	23.600	23.861m

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052722\
Data File : PL075126.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 May 2022 17:27
Operator : AR\AJ
Sample : PB145085BL
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
PB145085BL


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 05/31/2022
Supervised By : Ankita Jodhani 05/31/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 02:47:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051122.M
Quant Title : GC Extractables
QLast Update : Sat May 14 02:37:15 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.644 min
Delta R.T.: -0.004 min
Response: 11491107
Conc: 22.63 ng/ml

Instrument :
ECD_L
ClientSampleId :
PB145085BL


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 05/31/2022
Supervised By :Ankita Jodhani 05/31/2022




#1 Tetrachloro-m-xylene

R.T.: 4.644 min
Delta R.T.: -0.010 min
Response: 31915646
Conc: 23.25 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.639 min
Delta R.T.: 0.004 min
Response: 10539419
Conc: 23.60 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.390 min
Delta R.T.: -0.003 min
Response: 24491419
Conc: 23.86 ng/ml m