

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052820\
 Data File : PL058958.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 May 2020 17:17
 Operator : AJ\MA
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 18:17:33 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052020.M
 Quant Title : GC Extractables
 QLast Update : Wed May 20 11:27:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.529	4.077	94816681	137.2E6	51.710	50.644
2)	SA Decachlor...	8.309	9.188	82650278	111.7E6	48.918	42.371
Target Compounds							
2)	A alpha-BHC	3.971	4.470	144.9E6	204.2E6	53.287	50.885
3)	MA gamma-BHC...	4.257	4.757	126.4E6	194.1E6	53.559	51.222
4)	MA Heptachlor	4.563	5.271	119.3E6	200.4E6	53.057	52.319
5)	MB Aldrin	4.815	5.579	102.4E6	167.1E6	52.897	48.978
6)	B beta-BHC	4.506	4.925	52281487	84757232	50.870	48.710
7)	B delta-BHC	4.713	5.145	116.3E6	186.0E6	54.247	51.006
8)	B Heptachlo...	5.259	5.967	94976263	157.3E6	51.460	47.703
9)	A Endosulfan I	5.600	6.326	87539123	142.6E6	51.570	48.510
10)	B gamma-Chl...	5.488	6.201	94376220	157.2E6	51.477	47.449
11)	B alpha-Chl...	5.546	6.274	94029832	156.2E6	51.388	48.091
12)	B 4,4'-DDE	5.710	6.430	86057300	144.0E6	52.407	50.442
13)	MA Dieldrin	5.845	6.584	91698077	150.1E6	52.240	49.553
14)	MA Endrin	6.102	6.803	82787002	126.1E6	52.822	50.546
15)	B Endosulfa...	6.375	7.011	84683870	131.9E6	50.828	47.050
16)	A 4,4'-DDD	6.226	6.920	74683886	119.4E6	52.760	50.027
17)	MA 4,4'-DDT	6.469	7.221	76475275	117.2E6	52.574	49.280
18)	B Endrin al...	6.543	7.135	71259245	113.1E6	50.495	47.945
19)	B Endosulfa...	6.756	7.358	83315181	125.8E6	51.067	47.879
20)	A Methoxychlor	7.015	7.677	43951275	66461196	51.524	47.654
21)	B Endrin ke...	7.249	7.832	95577542	135.7E6	51.827	48.980
22)	Mirex	7.443	8.289	75109523	99446130	48.563	45.375

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052820\
Data File : PL058958.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2020 17:17
Operator : AJ\MA
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 18:17:33 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052020.M
Quant Title : GC Extractables
QLast Update : Wed May 20 11:27:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

