

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052820\
 Data File : PL058938.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 May 2020 10:10
 Operator : AJ\MA
 Sample : L2746-07
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 NM95-COMP-2020-0-6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 28 18:11:49 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052020.M
 Quant Title : GC Extractables
 QLast Update : Wed May 20 11:27:53 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.531	4.075	22874449	31804301	12.475	11.743
28) SA Decachlor...	8.311	9.188	17969594	24106841	10.636	9.145
Target Compounds						
12) B 4,4'-DDE	5.713	6.428	2850070	5668848	1.736	1.986
17) MA 4,4'-DDT	6.469	7.220	3714489	2172674	2.554	0.914 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052820\
Data File : PL058938.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2020 10:10
Operator : AJ\MA
Sample : L2746-07
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
NM95-COMP-2020-0-6

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 28 18:11:49 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052020.M
Quant Title : GC Extractables
QLast Update : Wed May 20 11:27:53 2020
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene


R.T.: 3.531 min
Delta R.T.: 0.000 min
Response: 22874449
Conc: 12.47 ng/ml

Instrument :
ECD_L
ClientSampleId :
NM95-COMP-2020-0-6




#1 Tetrachloro-m-xylene

R.T.: 4.075 min
Delta R.T.: -0.002 min
Response: 31804301
Conc: 11.74 ng/ml




#12 4,4'-DDE

R.T.: 5.713 min
Delta R.T.: 0.000 min
Response: 2850070
Conc: 1.74 ng/ml



#12 4,4'-DDE


R.T.: 6.428 min
Delta R.T.: 0.000 min
Response: 5668848
Conc: 1.99 ng/ml



#17 4,4'-DDT


R.T.: 6.469 min
Delta R.T.: -0.002 min
Response: 3714489
Conc: 2.55 ng/ml

Instrument :
ECD_L
ClientSampleId :
NM95-COMP-2020-0-6




#17 4,4'-DDT

R.T.: 7.220 min
Delta R.T.: 0.000 min
Response: 2172674
Conc: 0.91 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.311 min
Delta R.T.: -0.001 min
Response: 17969594
Conc: 10.64 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.188 min
Delta R.T.: -0.001 min
Response: 24106841
Conc: 9.14 ng/ml