

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052921\
 Data File : PL066952.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 May 2021 13:57
 Operator : AR\AJ
 Sample : PB136658BL
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 06:10:15 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052721CLP.M
 Quant Title : GC Extractables
 QLast Update : Thu May 27 02:30:58 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	4.482	5.533	11307291	17500387	21.188	19.791
2)	SA Decachlor...	10.229	11.437	29003703	37757415	43.254	41.357
Target Compounds							
7)	B delta-BHC	0.000	7.008	0	847636	N.D.	0.767
15)	B Endosulfa...	8.133f	0.000	3609551	0	6.011	N.D. #
16)	A 4,4'-DDD	0.000	9.034	0	5016834	N.D.	6.771

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL052921\
Data File : PL066952.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 May 2021 13:57
Operator : AR\AJ
Sample : PB136658BL
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 06:10:15 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052721CLP.M
Quant Title : GC Extractables
QLast Update : Thu May 27 02:30:58 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm




Instrument :
ECD_L
ClientSampled :



#1 Tetrachloro-m-xylene


R.T.: 4.482 min
Delta R.T.: 0.000 min
Response: 11307291
Conc: 21.19 ng/ml

Instrument :
ECD_L
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 5.533 min
Delta R.T.: 0.001 min
Response: 17500387
Conc: 19.79 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 6.240 min
Response: 0
Conc: N.D.



#7 delta-BHC


R.T.: 7.008 min
Delta R.T.: -0.020 min
Response: 847636
Conc: 0.77 ng/ml



#15 Endosulfan II


R.T.: 8.133 min
Delta R.T.: -0.056 min
Response: 3609551
Conc: 6.01 ng/ml

Instrument :
ECD_L
ClientSampleId :




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 9.158 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T. : 8.053 min
Response: 0
Conc: N.D.



#16 4,4'-DDD


R.T.: 9.034 min
Delta R.T.: -0.035 min
Response: 5016834
Conc: 6.77 ng/ml



#27 Decachlorobiphenyl

R.T.: 10.229 min
Delta R.T.: 0.003 min
Response: 29003703
Conc: 43.25 ng/ml

Instrument :
ECD_L
ClientSampleId :



#27 Decachlorobiphenyl

R.T.: 11.437 min
Delta R.T.: 0.005 min
Response: 37757415
Conc: 41.36 ng/ml