

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053023\
 Data File : PL083093.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2023 15:22
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 30 21:00:46 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.316	2.636	137.4E6	123.9E6	46.756	47.713
2)	SA Decachlor...	8.733	7.741	106.2E6	123.5E6	50.654	49.664
Target Compounds							
2)	A alpha-BHC	3.765	3.132	210.4E6	178.4E6	47.406	47.798
3)	MA gamma-BHC...	4.093	3.458	195.5E6	163.1E6	47.184	48.927
4)	MA Heptachlor	4.671	3.789	183.2E6	167.9E6	47.922	45.869
5)	MB Aldrin	5.008	4.065	171.7E6	154.8E6	48.802	47.564
6)	B beta-BHC	4.298	3.759	86516219	67767957	45.953	47.387
7)	B delta-BHC	4.538	3.984	191.6E6	159.4E6	55.708	48.655
8)	B Heptachlo...	5.434	4.566	157.1E6	142.2E6	49.115	47.909
9)	A Endosulfan I	5.815	4.933	145.7E6	133.7E6	48.560	47.839
10)	B gamma-Chl...	5.689	4.815	159.9E6	144.3E6	50.321	48.052
11)	B alpha-Chl...	5.767	4.877	156.8E6	144.1E6	48.861	48.018
12)	B 4,4'-DDE	5.946	5.068	138.5E6	123.9E6	50.230	51.010
13)	MA Dieldrin	6.089	5.197	147.6E6	135.8E6	49.070	48.104
14)	MA Endrin	6.317	5.471	126.6E6	119.1E6	47.126	46.965
15)	B Endosulfa...	6.539	5.768	120.4E6	114.1E6	45.124	46.134
16)	A 4,4'-DDD	6.462	5.623	113.7E6	100.7E6	52.303	52.361
17)	MA 4,4'-DDT	6.773	5.871	116.8E6	105.2E6	46.646	47.142
18)	B Endrin al...	6.670	5.948	95758501	91396209	47.626	46.219
19)	B Endosulfa...	6.904	6.172	125.5E6	117.1E6	48.964	47.828
20)	A Methoxychlor	7.253	6.450	62777310	59107510	47.473	46.477
21)	B Endrin ke...	7.383	6.676	129.6E6	132.9E6	50.804	49.136
22)	Mirex	7.847	6.846	100.7E6	111.5E6	48.043	47.668

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053023\
Data File : PL083093.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2023 15:22
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 30 21:00:46 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

