

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053119\
 Data File : PL049186.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2019 02:17
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 6/3/2019 2:17:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 04:00:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.324	3.937	593.1E6	157.0E6	53.825	54.355
28)	SA Decachlor...	7.992	8.973	625.4E6	177.6E6	49.185	53.191
Target Compounds							
2)	A alpha-BHC	3.752	4.325	988.9E6	228.8E6	54.886	57.092
3)	MA gamma-BHC...	4.028	4.608	923.5E6	212.4E6	54.486	56.724
4)	MA Heptachlor	4.313	5.111	968.8E6	221.0E6	54.707	54.995
5)	MB Aldrin	4.551	5.414	941.8E6	206.5E6	55.158	56.135
6)	B beta-BHC	4.282	4.783	379.6E6	95493172	54.624	55.440
7)	B delta-BHC	4.476	4.997	905.0E6	192.3E6	55.998	62.687
8)	B Heptachlo...	4.988	5.800	828.5E6	190.8E6	55.862	58.381m
9)	A Endosulfan I	5.319	6.156	776.5E6	175.7E6	53.058	54.537
10)	B gamma-Chl...	5.211	6.035	862.1E6	191.7E6	54.001	55.411
11)	B alpha-Chl...	5.267	6.106	827.2E6	189.0E6	52.952	54.481
12)	B 4,4'-DDE	5.434	6.266	841.4E6	179.0E6	53.789	54.528
13)	MA Dieldrin	5.558	6.412	835.1E6	184.4E6	54.615	54.986
14)	MA Endrin	5.811	6.630	752.8E6	160.9E6	52.196	53.919
15)	B Endosulfa...	6.086	6.840	691.0E6	159.3E6	52.328	52.532
16)	A 4,4'-DDD	5.946	6.757	641.1E6	143.8E6	53.869	54.687
17)	MA 4,4'-DDT	6.180	7.053	649.7E6	149.3E6	52.766	53.239
18)	B Endrin al...	6.253	6.965	568.7E6	137.2E6	51.961	53.183
19)	B Endosulfa...	6.464	7.188	671.9E6	159.5E6	51.754	52.469
20)	A Methoxychlor	6.725	7.514	328.7E6	86294320	51.247	52.052
21)	B Endrin ke...	6.946	7.656	755.9E6	165.2E6	55.200	51.928
22)	Mirex	7.120	8.106	517.4E6	130.2E6	49.106	50.119

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053119\
 Data File : PL049186.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 01 Jun 2019 02:17
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampled :
 PSTDCCC050

Manual Integrations
 APPROVED

Ankita
 6/3/2019 2:17:43 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 01 04:00:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

