

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB053123\  
 Data File : FB029901.D  
 Signal(s) : FID2B.CH  
 Acq On : 31 May 2023 11:14  
 Operator : YP/AJ  
 Sample : BSF0531S1  
 Misc : 5.00G/5.00 ML DI WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 ECD\_L  
 ClientSampleId :

Integration File: Calibration.e  
 Quant Time: Jun 01 09:08:18 2023  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB051123.M  
 Quant Title :  
 QLast Update : Thu May 11 11:33:20 2023  
 Response via : Initial Calibration  
 Integrator: ChemStation

Volume Inj. : 5 g/ml  
 Signal Phase : RTX-502.2  
 Signal Info : 60mx0.53mmx3.00um

| Compound                     | R.T.   | Response | Conc Units   |
|------------------------------|--------|----------|--------------|
| -----                        |        |          |              |
| System Monitoring Compounds  |        |          |              |
| 5) s AAA-TFT                 | 8.783  | 208174   | 17.095 ng/ml |
| Target Compounds             |        |          |              |
| 1) t 2-Methylpentane         | 4.700  | 448567   | 26.180 ng/ml |
| 2) t 2,2,4-Trimethylpentane  | 7.407  | 445593   | 26.108 ng/ml |
| 3) t n-Heptane               | 7.744  | 147547   | 8.309 ng/ml  |
| 4) t Benzene                 | 7.880  | 190975   | 9.116 ng/ml  |
| 6) t Toluene                 | 10.612 | 557799   | 27.395 ng/ml |
| 7) t Ethylbenzene            | 13.048 | 181062   | 9.448 ng/ml  |
| 8) t m-Xylene                | 13.180 | 358031   | 18.582 ng/ml |
| 9) t O-Xylene                | 13.907 | 347078   | 18.658 ng/ml |
| 10) t 1,2,4-Trimethylbenzene | 16.181 | 356253   | 19.152 ng/ml |
| -----                        |        |          |              |

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_B\Data\FB053123\  
Data File : FB029901.D  
Signal(s) : FID2B.CH  
Acq On : 31 May 2023 11:14  
Operator : YP/AJ  
Sample : BSF0531S1  
Misc : 5.00G/5.00 ML DI WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
ECD\_L  
ClientSampleId :

Integration File: Calibration.e  
Quant Time: Jun 01 09:08:18 2023  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_B\Method\FB051123.M  
Quant Title :  
QLast Update : Thu May 11 11:33:20 2023  
Response via : Initial Calibration  
Integrator: ChemStation

Volume Inj. : 5 g/ml  
Signal Phase : RTX-502.2  
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.700 min  
Delta R.T.: 0.000 min  
Response: 448567  
Conc: 26.18 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.407 min  
Delta R.T.: 0.000 min  
Response: 445593  
Conc: 26.11 ng/ml




#3 n-Heptane

R.T.: 7.744 min  
Delta R.T.: 0.002 min  
Response: 147547  
Conc: 8.31 ng/ml



#4 Benzene


R.T.: 7.880 min  
Delta R.T.: 0.002 min  
Response: 190975  
Conc: 9.12 ng/ml



#5 AAA-TFT


R.T.: 8.783 min  
Delta R.T.: 0.000 min  
Response: 208174  
Conc: 17.09 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :




#6 Toluene

R.T.: 10.612 min  
Delta R.T.: 0.001 min  
Response: 557799  
Conc: 27.39 ng/ml




#7 Ethylbenzene

R.T.: 13.048 min  
Delta R.T.: 0.001 min  
Response: 181062  
Conc: 9.45 ng/ml



#8 m-Xylene


R.T.: 13.180 min  
Delta R.T.: 0.001 min  
Response: 358031  
Conc: 18.58 ng/ml



#9 O-Xylene

R.T.: 13.907 min  
Delta R.T.: 0.001 min  
Response: 347078  
Conc: 18.66 ng/ml

Instrument :  
ECD\_L  
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.181 min  
Delta R.T.: 0.000 min  
Response: 356253  
Conc: 19.15 ng/ml