

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053124\
 Data File : PL089828.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 May 2024 16:30
 Operator : AR\AJ
 Sample : HEXANE
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 31 00:21:47 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051724CLP.M
 Quant Title : GC Extractables
 QLast Update : Fri May 17 15:59:18 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds							
1)	SA Tetrachlo...	3.585f	0.000	446900	0	0.230	N.D. #
Target Compounds							
2)	A alpha-BHC	4.006	0.000	163515	0	<MDL	N.D. #
8)	B Heptachlo...	5.691	0.000	9514404	0	4.114	N.D. #
9)	A Endosulfan I	6.112	0.000	1893937	0	0.944	N.D. #
12)	B 4,4'-DDE	6.264f	5.210f	3640443	1194893	1.989	0.851 #
14)	MA Endrin	6.566	5.681	8485964	1559073	4.982	1.132 #
15)	B Endosulfa...	6.824	0.000	720957	0	0.399	N.D. #
16)	A 4,4'-DDD	6.773f	0.000	410943	0	0.279	N.D. #
17)	MA 4,4'-DDT	7.036	0.000	241823	0	0.172	N.D. #
18)	B Endrin al...	6.982f	0.000	603889	0	0.443	N.D. #
22)	Toxaphene-1	5.852	0.000	3400062	0	458.871	N.D. #
23)	Toxaphene-2	6.506f	0.000	10288938	0	517.622	N.D. #
24)	Toxaphene-3	7.036	0.000	241823	0	5.546	N.D. #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL053124\
Data File : PL089828.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 May 2024 16:30
Operator : AR\AJ
Sample : HEXANE
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 31 00:21:47 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051724CLP.M
Quant Title : GC Extractables
QLast Update : Fri May 17 15:59:18 2024
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm






#1 Tetrachloro-m-xylene

R.T.: 3.585 min
Delta R.T.: 0.035 min
Response: 446900
Conc: 0.23 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 0.000 min
Exp R.T.: 2.795 min
Response: 0
Conc: N.D.




#8 Heptachlor epoxide

R.T.: 5.691 min
Delta R.T.: -0.004 min
Response: 9514404
Conc: 4.11 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T.: 4.757 min
Response: 0
Conc: N.D.




#9 Endosulfan I

R.T.: 6.112 min
Delta R.T.: 0.032 min
Response: 1893937
Conc: 0.94 ng/ml




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T.: 5.128 min
Response: 0
Conc: N.D.




#12 4,4'-DDE

R.T.: 6.264 min
Delta R.T.: 0.062 min
Response: 3640443
Conc: 1.99 ng/ml




#12 4,4'-DDE

R.T.: 5.210 min
Delta R.T.: -0.050 min
Response: 1194893
Conc: 0.85 ng/ml




#14 Endrin

R.T.: 6.566 min
Delta R.T.: -0.019 min
Response: 8485964
Conc: 4.98 ng/ml




#14 Endrin

R.T.: 5.681 min
Delta R.T.: 0.013 min
Response: 1559073
Conc: 1.13 ng/ml




#15 Endosulfan II

R.T.: 6.824 min
Delta R.T.: 0.021 min
Response: 720957
Conc: 0.40 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T.: 5.961 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 6.773 min
Delta R.T.: 0.055 min
Response: 410943
Conc: 0.28 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T.: 5.815 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 7.036 min
Delta R.T.: 0.004 min
Response: 241823
Conc: 0.17 ng/ml




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T.: 6.066 min
Response: 0
Conc: N.D.




#18 Endrin aldehyde

R.T.: 6.982 min
Delta R.T.: 0.050 min
Response: 603889
Conc: 0.44 ng/ml




#18 Endrin aldehyde

R.T.: 0.000 min
Exp R.T.: 6.141 min
Response: 0
Conc: N.D.




#22 Toxaphene-1

R.T.: 5.852 min
Delta R.T.: -0.010 min
Response: 3400062
Conc: 458.87 ng/ml




#22 Toxaphene-1

R.T.: 0.000 min
Exp R.T.: 5.032 min
Response: 0
Conc: N.D.




#23 Toxaphene-2

R.T.: 6.506 min
Delta R.T.: 0.055 min
Response: 10288938
Conc: 517.62 ng/ml




#23 Toxaphene-2

R.T.: 0.000 min
Exp R.T. : 5.357 min
Response: 0
Conc: N.D.




#24 Toxaphene-3

R.T.: 7.036 min
Delta R.T.: -0.032 min
Response: 241823
Conc: 5.55 ng/ml




#24 Toxaphene-3

R.T.: 0.000 min
Exp R.T. : 6.631 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 9.065 min
Response: 0
Conc: N.D.



#27 Decachlorobiphenyl

R.T.: 0.000 min
Exp R.T. : 7.946 min
Response: 0
Conc: N.D.