

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060222\
 Data File : PL075403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2022 19:35
 Operator : AR\AJ
 Sample : N2933-04
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 06:08:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052822.M
 Quant Title : GC Extractables
 QLast Update : Wed Jun 01 10:19:11 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml



System Monitoring Compounds							
1)	SA Tetrachlo...	5.644	4.642	10486387	28975030	17.842	18.291
28)	SA Decachlor...	11.640	10.387	7719264	19254762	14.397	14.031
Target Compounds							
2)	A alpha-BHC	6.201f	0.000	327713	0	0.397	N.D. #
5)	MB Aldrin	0.000	6.495	0	1082654	N.D.	0.561 #
7)	B delta-BHC	0.000	6.412	0	87602	N.D.	0.041 #
8)	B Heptachlo...	8.126	0.000	445879	0	0.653	N.D. #
9)	A Endosulfan I	0.000	7.471	0	156511	N.D.	0.093 #
11)	B alpha-Chl...	0.000	7.394f	0	337466	N.D.	0.185 #
13)	MA Dieldrin	8.808	0.000	226654	0	0.358	N.D. #
14)	MA Endrin	0.000	8.067f	0	559842	N.D.	0.360 #
15)	B Endosulfa...	9.307	0.000	306183	0	0.542	N.D. #
16)	A 4,4'-DDD	9.204	0.000	493669	0	0.998	N.D. #
17)	MA 4,4'-DDT	0.000	8.464	0	1665893	N.D.	1.165 #
19)	B Endosulfa...	0.000	8.775	0	576944	N.D.	0.385 #
21)	B Endrin ke...	10.187f	9.327f	-101549	649612	N.D.	0.406
24)	Chlordane-2	0.000	6.642	0	3977791	N.D.	58.223 #
26)	Chlordane-4	0.000	7.394f	0	337466	N.D.	1.786 #
27)	Chlordane-5	0.000	8.128	0	1232443	N.D.	24.698 #


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060222\
Data File : PL075403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2022 19:35
Operator : AR\AJ
Sample : N2933-04
Misc :
ALS Vial : 34 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 06:08:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052822.M
Quant Title : GC Extractables
QLast Update : Wed Jun 01 10:19:11 2022
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.644 min
Delta R.T.: 0.003 min
Response: 10486387
Conc: 17.84 ng/ml




#1 Tetrachloro-m-xylene

R.T.: 4.642 min
Delta R.T.: 0.000 min
Response: 28975030
Conc: 18.29 ng/ml




#2 alpha-BHC

R.T.: 6.201 min
Delta R.T.: -0.024 min
Response: 327713
Conc: 0.40 ng/ml




#2 alpha-BHC

R.T.: 0.000 min
Exp R.T.: 5.344 min
Response: 0
Conc: N.D.




#5 Aldrin

R.T.: 0.000 min
Exp R.T. : 7.654 min
Response: 0
Conc: N.D.




#5 Aldrin

R.T.: 6.495 min
Delta R.T.: -0.003 min
Response: 1082654
Conc: 0.56 ng/ml




#7 delta-BHC

R.T.: 0.000 min
Exp R.T. : 7.146 min
Response: 0
Conc: N.D.




#7 delta-BHC

R.T.: 6.412 min
Delta R.T.: 0.009 min
Response: 87602
Conc: 0.04 ng/ml




#8 Heptachlor epoxide

R.T.: 8.126 min
Delta R.T.: 0.005 min
Response: 445879
Conc: 0.65 ng/ml




#8 Heptachlor epoxide

R.T.: 0.000 min
Exp R.T. : 7.072 min
Response: 0
Conc: N.D.




#9 Endosulfan I

R.T.: 0.000 min
Exp R.T. : 8.530 min
Response: 0
Conc: N.D.




#9 Endosulfan I

R.T.: 7.471 min
Delta R.T.: -0.004 min
Response: 156511
Conc: 0.09 ng/ml




#11 alpha-Chlordane

R.T.: 0.000 min
Exp R.T. : 8.474 min
Response: 0
Conc: N.D.




#11 alpha-Chlordane

R.T.: 7.394 min
Delta R.T.: -0.024 min
Response: 337466
Conc: 0.18 ng/ml




#13 Dieldrin

R.T.: 8.808 min
Delta R.T.: -0.013 min
Response: 226654
Conc: 0.36 ng/ml




#13 Dieldrin

R.T.: 0.000 min
Exp R.T. : 7.760 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 0.000 min
Exp R.T. : 9.062 min
Response: 0
Conc: N.D.




#14 Endrin

R.T.: 8.067 min
Delta R.T.: 0.016 min
Response: 559842
Conc: 0.36 ng/ml




#15 Endosulfan II

R.T.: 9.307 min
Delta R.T.: 0.011 min
Response: 306183
Conc: 0.54 ng/ml




#15 Endosulfan II

R.T.: 0.000 min
Exp R.T. : 8.361 min
Response: 0
Conc: N.D.




#16 4,4'-DDD

R.T.: 9.204 min
Delta R.T.: 0.003 min
Response: 493669
Conc: 1.00 ng/ml




#16 4,4'-DDD

R.T.: 0.000 min
Exp R.T.: 8.215 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 0.000 min
Exp R.T.: 9.522 min
Response: 0
Conc: N.D.




#17 4,4'-DDT

R.T.: 8.464 min
Delta R.T.: -0.011 min
Response: 1665893
Conc: 1.16 ng/ml




#19 Endosulfan Sulfate

R.T.: 0.000 min
Exp R.T.: 9.669 min
Response: 0
Conc: N.D.




#19 Endosulfan Sulfate

R.T.: 8.775 min
Delta R.T.: -0.004 min
Response: 576944
Conc: 0.39 ng/ml




#21 Endrin ketone

R.T.: 10.187 min
Delta R.T.: 0.016 min
Response: -101549
Conc: N.D.




#21 Endrin ketone

R.T.: 9.327 min
Delta R.T.: 0.029 min
Response: 649612
Conc: 0.41 ng/ml




#24 Chlordane-2

R.T.: 0.000 min
Exp R.T.: 7.638 min
Response: 0
Conc: N.D.




#24 Chlordane-2

R.T.: 6.642 min
Delta R.T.: -0.011 min
Response: 3977791
Conc: 58.22 ng/ml




#26 Chlordane-4

R.T.: 0.000 min
Exp R.T.: 8.482 min
Response: 0
Conc: N.D.




#26 Chlordane-4

R.T.: 7.394 min
Delta R.T.: -0.027 min
Response: 337466
Conc: 1.79 ng/ml




#27 Chlordane-5

R.T.: 0.000 min
Exp R.T.: 9.375 min
Response: 0
Conc: N.D.




#27 Chlordane-5

R.T.: 8.128 min
Delta R.T.: -0.005 min
Response: 1232443
Conc: 24.70 ng/ml



#28 Decachlorobiphenyl

R.T.: 11.640 min
Delta R.T.: 0.000 min
Response: 7719264
Conc: 14.40 ng/ml



#28 Decachlorobiphenyl

R.T.: 10.387 min
Delta R.T.: -0.005 min
Response: 19254762
Conc: 14.03 ng/ml