

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060225\
 Data File : PL095870.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2025 15:58
 Operator : AR\AJ
 Sample : Q2160-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TP05-MHH-WC

Manual Integrations
 APPROVED

Reviewed By :Abdul Mirza 06/03/2025
 Supervised By :mohammad ahmed 06/04/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 21:53:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
 Quant Title : GC Extractables
 QLast Update : Thu May 22 06:29:30 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.571	2.886	60263285	72502931	19.099m	18.524
28) SA Decachlor...	9.100	8.061	42218770	72503649	17.917	16.574

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060225\
Data File : PL095870.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2025 15:58
Operator : AR\AJ
Sample : Q2160-05
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
TP05-MHH-WC


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/03/2025
Supervised By : mohammad ahmed 06/04/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 21:53:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052125.M
Quant Title : GC Extractables
QLast Update : Thu May 22 06:29:30 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.571 min
Delta R.T.: -0.006 min
Response: 60263285
Conc: 19.10 ng/ml

Instrument :
ECD_L
ClientSampleId :
TP05-MHH-WC


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/03/2025
Supervised By :mohammad ahmed 06/04/2025




#1 Tetrachloro-m-xylene

R.T.: 2.886 min
Delta R.T.: 0.002 min
Response: 72502931
Conc: 18.52 ng/ml



#28 Decachlorobiphenyl

R.T.: 9.100 min
Delta R.T.: -0.009 min
Response: 42218770
Conc: 17.92 ng/ml



#28 Decachlorobiphenyl

R.T.: 8.061 min
Delta R.T.: -0.003 min
Response: 72503649
Conc: 16.57 ng/ml