

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060319\
 Data File : PL049242.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2019 18:21
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 04 01:55:54 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.324	3.936	545.5E6	144.4E6	49.510	49.985
2)	SA Decachlor...	7.991	8.971	566.0E6	155.5E6	44.511	46.551
Target Compounds							
2)	A alpha-BHC	3.751	4.324	922.5E6	211.5E6	51.205	52.770
3)	MA gamma-BHC...	4.027	4.608	853.6E6	195.7E6	50.363	52.286
4)	MA Heptachlor	4.312	5.111	887.8E6	202.6E6	50.133	50.408
5)	MB Aldrin	4.550	5.413	843.3E6	184.7E6	49.387	50.188
6)	B beta-BHC	4.281	4.782	351.6E6	88870292	50.600	51.595
7)	B delta-BHC	4.475	4.997	871.5E6	179.0E6	53.926	58.346
8)	B Heptachlo...	4.987	5.800	770.0E6	171.2E6	51.919	52.388
9)	A Endosulfan I	5.318	6.155	712.9E6	159.6E6	48.711	49.556
10)	B gamma-Chl...	5.209	6.034	802.1E6	176.5E6	50.245	51.014
11)	B alpha-Chl...	5.265	6.105	771.4E6	175.0E6	49.377	50.433
12)	B 4,4'-DDE	5.433	6.265	765.5E6	164.8E6	48.939	50.214
13)	MA Dieldrin	5.557	6.411	769.4E6	169.6E6	50.316	50.560
14)	MA Endrin	5.809	6.628	693.6E6	146.5E6	48.088	49.100
15)	B Endosulfa...	6.085	6.840	653.9E6	150.0E6	49.516	49.463
16)	A 4,4'-DDD	5.944	6.755	588.5E6	133.0E6	49.453	50.557
17)	MA 4,4'-DDT	6.178	7.052	590.5E6	134.9E6	47.963	48.101
18)	B Endrin al...	6.251	6.964	536.9E6	129.2E6	49.062	50.103
19)	B Endosulfa...	6.463	7.186	595.1E6	142.1E6	45.835	46.730
20)	A Methoxychlor	6.724	7.513	296.9E6	76543787	46.288	46.171
21)	B Endrin ke...	6.944	7.655	670.4E6	154.1E6	48.960	48.455
22)	Mirex	7.119	8.105	468.4E6	120.3E6	44.449	46.300

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060319\
Data File : PL049242.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2019 18:21
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 04 01:55:54 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
Quant Title : GC Extractables
QLast Update : Fri May 31 01:35:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

