

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060324\
 Data File : PL089891.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 Jun 2024 09:39
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 03 23:10:38 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052424.M
 Quant Title : GC Extractables
 QLast Update : Fri May 24 14:57:01 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.553	2.792	109.8E6	81466539	52.421	56.817
28)	SA Decachlor...	9.068	7.942	71239985	74403072	48.450	50.069
Target Compounds							
2)	A alpha-BHC	4.009	3.297	154.8E6	121.3E6	51.923	56.763
3)	MA gamma-BHC...	4.341	3.628	150.6E6	117.7E6	51.810	56.268
4)	MA Heptachlor	4.931	3.969	131.8E6	110.3E6	50.502	55.260
5)	MB Aldrin	5.273	4.250	132.3E6	112.1E6	49.783	56.118
6)	B beta-BHC	4.537	3.925	66641855	50215872	51.021	54.892
7)	B delta-BHC	4.784	4.156	142.8E6	115.1E6	51.849	57.120
8)	B Heptachlo...	5.698	4.753	121.2E6	100.2E6	48.668	54.612
9)	A Endosulfan I	6.083	5.123	109.9E6	93569790	48.682	54.771
10)	B gamma-Chl...	5.954	5.003	115.9E6	101.7E6	47.943	54.589
11)	B alpha-Chl...	6.033	5.067	117.0E6	101.7E6	48.703	54.181
12)	B 4,4'-DDE	6.205	5.256	105.0E6	92028777	48.260	55.085
13)	MA Dieldrin	6.358	5.388	115.9E6	99830217	48.892	54.178
14)	MA Endrin	6.587	5.663	95656161	84571826	47.579	51.632
15)	B Endosulfa...	6.805	5.958	100.5E6	85564405	48.197	53.130
16)	A 4,4'-DDD	6.720	5.811	89080584	79147218	50.856	56.149
17)	MA 4,4'-DDT	7.035	6.062	86025046	76986994	46.725	50.696
18)	B Endrin al...	6.935	6.137	73932027	62965184	47.696	51.885
19)	B Endosulfa...	7.169	6.360	92380351	83715176	48.271	53.231
20)	A Methoxychlor	7.510	6.638	46410542	42937513	47.253	48.267
21)	B Endrin ke...	7.654	6.865	103.0E6	99083729	48.915	54.372
22)	Mirex	8.129	7.048	76747430	77644943	46.133	50.558

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060324\
Data File : PL089891.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 Jun 2024 09:39
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 03 23:10:38 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL052424.M
Quant Title : GC Extractables
QLast Update : Fri May 24 14:57:01 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

