

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060419\
 Data File : PL049298.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2019 02:00
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 05 04:12:17 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.324	3.937	522.7E6	143.1E6	47.440	49.517
28)	SA Decachlor...	7.991	8.972	479.0E6	142.3E6	37.673	42.617
Target Compounds							
2)	A alpha-BHC	3.751	4.324	880.4E6	211.0E6	48.867	52.642
3)	MA gamma-BHC...	4.027	4.608	814.0E6	194.6E6	48.027	51.983
4)	MA Heptachlor	4.312	5.111	820.7E6	194.7E6	46.345	48.451
5)	MB Aldrin	4.550	5.413	793.4E6	181.5E6	46.465	49.324
6)	B beta-BHC	4.281	4.783	336.6E6	88318653	48.443	51.275
7)	B delta-BHC	4.475	4.997	824.9E6	185.7E6	51.043	60.559
8)	B Heptachlor...	4.987	5.800	727.4E6	183.7E6	49.050	56.199
9)	A Endosulfan I	5.318	6.155	668.7E6	147.3E6	45.690	45.723
10)	B gamma-Chl...	5.210	6.034	757.4E6	165.7E6	47.445	47.890
11)	B alpha-Chl...	5.266	6.106	725.1E6	161.7E6	46.414	46.601
12)	B 4,4'-DDE	5.434	6.266	733.5E6	146.6E6	46.894	44.656
13)	MA Dieldrin	5.557	6.412	724.7E6	152.6E6	47.395	45.494
14)	MA Endrin	5.810	6.628	620.2E6	128.0E6	42.997	42.878
15)	B Endosulfa...	6.085	6.840	564.7E6	131.8E6	42.765	43.469
16)	A 4,4'-DDD	5.945	6.756	567.1E6	120.6E6	47.650	45.845
17)	MA 4,4'-DDT	6.179	7.053	431.1E6	102.4E6	35.013	36.531
18)	B Endrin al...	6.252	6.964	432.1E6	112.4E6	39.478	43.590
19)	B Endosulfa...	6.463	7.187	456.1E6	119.5E6	35.132	39.285
20)	A Methoxychlor	6.724	7.513	233.9E6	67099780	36.469	40.474
21)	B Endrin ke...	6.945	7.655	522.1E6	126.0E6	36.682	39.611
22)	Mirex	7.119	8.105	322.0E6	91631922	30.558	35.277

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060419\
Data File : PL049298.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2019 02:00
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 05 04:12:17 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
Quant Title : GC Extractables
QLast Update : Fri May 31 01:35:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

