

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060422\
 Data File : PL075538.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 Jun 2022 22:36
 Operator : AR\AJ
 Sample : N2897-03
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 P041-SS001-1218-01

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/06/2022
 Supervised By : Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 00:21:08 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060422.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 06 00:15:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	5.643	4.642	2507228	6726887	4.063	4.380
2)	SA Decachlor...	11.639	10.386	7618923	18443331	14.673	15.185
Target Compounds							
12)	B 4,4'-DDE	8.658	7.626	366364	945229	0.624m	0.665

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060422\
Data File : PL075538.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 Jun 2022 22:36
Operator : AR\AJ
Sample : N2897-03
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_L
ClientSampleId :
P041-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 06/06/2022
Supervised By : Ankita Jodhani 06/06/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 00:21:08 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060422.M
Quant Title : GC Extractables
QLast Update : Mon Jun 06 00:15:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm






#1 Tetrachloro-m-xylene

R.T.: 5.643 min
Delta R.T.: 0.000 min
Response: 2507228
Conc: 4.06 ng/ml

Instrument :
ECD_L
ClientSampleId :
P041-SS001-1218-01


Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/06/2022
Supervised By :Ankita Jodhani 06/06/2022




#1 Tetrachloro-m-xylene

R.T.: 4.642 min
Delta R.T.: 0.000 min
Response: 6726887
Conc: 4.38 ng/ml




#12 4,4'-DDE

R.T.: 8.658 min
Delta R.T.: -0.003 min
Response: 366364
Conc: 0.62 ng/ml m



#12 4,4'-DDE

R.T.: 7.626 min
Delta R.T.: -0.002 min
Response: 945229
Conc: 0.66 ng/ml




#28 Decachlorobiphenyl

R.T.: 11.639 min
Delta R.T.: -0.002 min
Response: 7618923
Conc: 14.67 ng/ml

Instrument :
ECD_L
ClientSampleId :
P041-SS001-1218-01

Manual Integrations
APPROVED

Reviewed By :Abdul Mirza 06/06/2022
Supervised By :Ankita Jodhani 06/06/2022



#28 Decachlorobiphenyl

R.T.: 10.386 min
Delta R.T.: 0.000 min
Response: 18443331
Conc: 15.19 ng/ml