

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060521\
 Data File : PL067095.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2021 11:39
 Operator : AR\AJ
 Sample : TOXAPH401
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 TOXAPH401

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 02:01:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060521CLP.M
 Quant Title : GC Extractables
 QLast Update : Mon Jun 07 01:59:57 2021
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.016	4.901	18973814	31405683	39.280	39.459
27)	SA Decachlor...	9.237	10.441	44300406	62830927	75.175	75.315
Target Compounds							
22)	Toxaphene-1	6.664	7.640	13118691	16823613	3956.973	3758.543
23)	Toxaphene-2	6.827	7.949	14528806	19580407	3842.760	4029.365
24)	Toxaphene-3	7.294	8.374	42766790	25273650	3940.545	3863.223
25)	Toxaphene-4	7.549	8.454	46023639	51319889	3936.144	4120.797
26)	Toxaphene-5	8.048	8.975	49437327	39172647	3955.254	3910.772

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060521\
Data File : PL067095.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2021 11:39
Operator : AR\AJ
Sample : TOXAPH401
Misc :
ALS Vial : 7 Sample Multiplier: 1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 02:01:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL060521CLP.M
Quant Title : GC Extractables
QLast Update : Mon Jun 07 01:59:57 2021
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x 0.2 Signal #2 Info : 30M x 0.32mm x 0.50µm

