

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060523\
 Data File : PL083235.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2023 19:35
 Operator : AR\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 06/06/2023
 Supervised By : Ankita Jodhani 06/06/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 05 21:19:55 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
 Quant Title : GC Extractables
 QLast Update : Sat May 20 04:58:53 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.314	2.632	134.2E6	120.7E6	45.676	46.495
28)	SA Decachlor...	8.730	7.736	98810605	116.2E6	47.113	46.714
Target Compounds							
2)	A alpha-BHC	3.763	3.129	205.1E6	175.6E6	46.193	47.060
3)	MA gamma-BHC...	4.091	3.455	187.5E6	161.6E6	45.259	48.471
4)	MA Heptachlor	4.668	3.786	172.9E6	162.6E6	45.244	44.420
5)	MB Aldrin	5.005	4.062	168.9E6	156.8E6	47.995	48.191
6)	B beta-BHC	4.295	3.756	84051105	67903087	44.643	47.482
7)	B delta-BHC	4.536	3.981	186.8E6	159.8E6	54.333	48.758
8)	B Heptachlo...	5.431	4.562	153.3E6	140.7E6	47.920	47.409
9)	A Endosulfan I	5.812	4.929	143.2E6	127.7E6	47.708	45.688
10)	B gamma-Chl...	5.685	4.811	151.7E6	143.6E6	47.741m	47.803
11)	B alpha-Chl...	5.764	4.874	153.4E6	142.9E6	47.794	47.597
12)	B 4,4'-DDE	5.942	5.063	131.1E6	122.4E6	47.560	50.414m
13)	MA Dieldrin	6.087	5.192	139.0E6	134.4E6	46.198	47.619m
14)	MA Endrin	6.313	5.467	121.7E6	124.1E6	45.284	48.927
15)	B Endosulfa...	6.537	5.765	117.1E6	111.5E6	43.743	45.072
16)	A 4,4'-DDD	6.459	5.619	109.2E6	99487422	50.254	51.729
17)	MA 4,4'-DDT	6.770	5.865	108.5E6	100.6E6	43.332	45.095m
18)	B Endrin al...	6.665	5.944	93500392	88379256	46.503m	44.693
19)	B Endosulfa...	6.901	6.167	122.4E6	115.1E6	47.755	46.982
20)	A Methoxychlor	7.252	6.446	57528262	55980204	43.504	44.018
21)	B Endrin ke...	7.381	6.671	123.8E6	126.0E6	48.516	46.591
22)	Mirex	7.843	6.840	95072712	107.0E6	45.358	45.739m

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060523\
Data File : PL083235.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2023 19:35
Operator : AR\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :

ECD_L

ClientSampleId :

PSTDCCC050

Manual Integrations

APPROVED

Reviewed By :Abdul Mirza 06/06/2023

Supervised By :Ankita Jodhani 06/06/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 05 21:19:55 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL051923.M
Quant Title : GC Extractables
QLast Update : Sat May 20 04:58:53 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

