

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060719\
 Data File : PL049368.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 07 Jun 2019 11:19
 Operator : SM\AJ
 Sample : PSTDCCC050
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_L
 ClientSampleId :
 PSTDCCC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 22:53:49 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
 Quant Title : GC Extractables
 QLast Update : Fri May 31 01:35:35 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
 Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.325	3.937	567.3E6	148.0E6	51.488	51.221
28)	SA Decachlor...	7.992	8.972	621.7E6	166.8E6	48.891	49.948
Target Compounds							
2)	A alpha-BHC	3.753	4.325	961.1E6	218.1E6	53.343	54.419
3)	MA gamma-BHC...	4.029	4.608	891.2E6	201.4E6	52.585	53.786
4)	MA Heptachlor	4.313	5.111	928.0E6	208.8E6	52.405	51.963
5)	MB Aldrin	4.552	5.414	869.1E6	188.1E6	50.900	51.135
6)	B beta-BHC	4.283	4.783	366.8E6	89026351	52.781	51.686
7)	B delta-BHC	4.477	4.997	906.0E6	184.0E6	56.063	59.997
8)	B Heptachlor...	4.989	5.800	744.6E6	188.3E6	50.210	57.600
9)	A Endosulfan I	5.319	6.155	745.9E6	161.8E6	50.968	50.238
10)	B gamma-Chl...	5.211	6.034	832.0E6	178.3E6	52.118	51.552
11)	B alpha-Chl...	5.267	6.106	799.8E6	175.3E6	51.194	50.533
12)	B 4,4'-DDE	5.435	6.266	826.1E6	164.8E6	52.814	50.189
13)	MA Dieldrin	5.558	6.412	801.3E6	170.1E6	52.404	50.719
14)	MA Endrin	5.811	6.629	763.0E6	155.8E6	52.900	52.221
15)	B Endosulfa...	6.086	6.841	689.5E6	148.2E6	52.210	48.844
16)	A 4,4'-DDD	5.946	6.757	629.8E6	130.4E6	52.920	49.573
17)	MA 4,4'-DDT	6.180	7.053	643.9E6	136.9E6	52.296	48.813
18)	B Endrin al...	6.253	6.964	566.6E6	126.6E6	51.770	49.078
19)	B Endosulfa...	6.464	7.187	603.9E6	139.9E6	46.511	46.021
20)	A Methoxychlor	6.725	7.514	304.0E6	81833371	47.404	49.361
21)	B Endrin ke...	6.946	7.655	695.5E6	146.2E6	48.871	45.959
22)	Mirex	7.120	8.106	460.4E6	114.0E6	43.692	43.886

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_L\Data\PL060719\
Data File : PL049368.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 07 Jun 2019 11:19
Operator : SM\AJ
Sample : PSTDCCC050
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_L
ClientSampleId :
PSTDCCC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 22:53:49 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_L\methods\PL053019.M
Quant Title : GC Extractables
QLast Update : Fri May 31 01:35:35 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR1
Signal #1 Info : 30M x 0.32mm x0.2 Signal #2 Info : 30M x 0.32mm x0.5µm

